



# Accumen Calibration Laboratory Pvt Ltd

(Calibration Division)

Plot No. 2, Araj No.102, 2nd Floor, Main Road Barra- 2, Kanpur -208027

Mob.:+91-8802463273, 7007115151, 8750869237, Ph. : 5123142732

E-mail:- accumencalibration22@gmail.com



CC-3727

Format No.:7.8F-01

Page 1 of 1

<b>ULR No. : CC372724000000146F</b>	<b>Certificate No.</b>	: ACL/000446
<b>CUSTOMER NAME &amp; ADDRESS:-</b> M/s. MD Path Lab 104/264, Rambagh, Kanpur- 208012 (U.P.) India	<b>Service Request No</b>	: 0045
	<b>Service Request Date.</b>	: 12-01-2024
	<b>Date of Receipt of UUC</b>	: 13-01-2024
	<b>Date of Calibration</b>	: 13-01-2024
	<b>Sugg. Due Date By Customer</b>	: 12-01-2025
	<b>Date of Issue</b>	: 14-01-2024

### ITEM'S DESCRIPTION & IDENTIFICATION

<b>Instrument Name</b>	: CENTRIFUGE	<b>Range/Size</b>	: Up to 5000 RPM
<b>Make</b>	: Med1	<b>Least Count</b>	: 1 RPM
<b>Model No.</b>	: ----	<b>Visual Inspection</b>	: Ok
<b>Serial No.</b>	: ----	<b>Location</b>	: Lab
<b>Party ID Mark</b>	: CF-01		

<b>ENVIRONMENTAL CONDITIONS</b>		<b>Calibration Procedure</b>	: MCP-23
<b>Temperature</b>	: 26.6°C	<b>Reference Standard</b>	: IS 12508 : 2013
<b>Relative Humidity</b>	: 58%	<b>Calibration Performed at</b>	: Site

<b>MAJOR EQUIPMENT USED</b>			
<b>Nomenclature</b>	<b>Range</b>	<b>Serial/Ident. No.</b>	<b>Traceability Certificate No.</b>
Digital tachometer (Contact /Non Contact)	up to 99999.9RPM	204-0014V/ACL/MMI/07	QCTS/00057/03
			<b>Valid upto</b> 04-04-2024

### CALIBRATION RESULT

Observed Value in UUC (in RPM)	Observed Value in Standard (in RPM)	Error ± (in RPM)	Uncertainty of Measurement (±)%
1000	999.9	0.1	3.1
2000	1999.9	0.1	3.1
3000	2998.9	1.1	3.1
5000	4998.9	1.1	3.1


Remarks:

Note :

- DUC = Device Under Calibration,UUC=Unit Under Calibration
- Expanded uncertainty of measurement at approx 95% confidence level for coverage factor(k)=2
- This Certificate refers only to the particular items submitted for calibration.
- Suggested next due date of calibration is given as requested by customer.
- DUC has been calibrated against lab standards are traceable to national & international standards.
- The result reported in this certificate are valid at the time of and under the stated conditions of Measurement.
- This certificate shall not be reproduced, except in full, without permission of CEO, Accumen Calibration Laboratory Pvt. Ltd., Kanpur (U.P.)

Issue by

Manish Verma  
(Calibrated By)



Jaysagar yadav  
( TM/QM )

\*End of Report\*