



Premium Calibration Services

PLOT No - 73, ADARSH NAGAR COLONY, MODINAGR ROAD, HAPUR-245101

Ph:- 0122-7965803, 08279757019

E-mail: Pcslab141@gmail.com



CC - 3740

CALIBRATION CERTIFICATE

ULR No. : CC37402400000049F	Calibration Done On : 20.01.2024	S.R.F.No : 0001/20
Issue Date : 21.01.2024	Calibration Valid Upto : 19.01.2025	S.R.F Date : 20.01.2024
UUC Receipt Date: ---	certificate No. PCS/0049F	Pages : 1 of 1

Calibrated for M/S.

Civil Hospital Ashta Central Processing Lab
Near Bus Stand Ashta MP-466116

Identification on UUC (Unit Under Calibration)

Instrument Name	Centrifuge	Size/Range	0 - 4000 RPM
Least Count	1 RPM	Conition Of Item	Good
Manufactured By	----	Calibration Performed at	Site
Manufacturer M/Sr.No.	---	Location	Lab
Party ID Mark	CH ASHTA/CPL/CENTR/09	Ref. Standard/Procedure	IS: 12508

Calibration Procedure / Method : As per Calibration Procedure Document No. PCS/CP/01, Section No.M/CP-21

Environmental conditions (s) : Temperature : 25 ± 5 °C. Relative Humidity : 30 to 80 % RH

Standard & Major Equipment(s)

Standard(s)& Equipments	Make	ID No. / Serial No.	Calibrating Agency	Traceability ULR No. / Cert. No.	Valid Up to
Dig. Tachometer	Line Seiki	PCS-TM-01 /	KL Technology Solutions	CC325723000000071F	21.02.2024

Uncertainty of measurement (at Approx 95% Confidence Level With Coverage Factor k=2) = ± 0.04 %

Calibration Results:

U.U.C Value (RPM)	Standard Value (RPM)	Error (RPM)
500	500.3	-0.3
1000	1000.1	-0.1
1500	1500.5	-0.5
3000	3000.8	-0.8
4000	4000.9	-0.9


***** End of Certificate *****

NOTE:

- The result are valid only for this Calibrated item.
- The calibration report shall not be reproduced without the written approval of calibration laboratory.
- Result reported are valid at the time of & under the stated conditions of measurement
- Laboratory standards are traceable to National/International Standards.
- Calibration Certificates issued for Weight & Measure parameters i.e. Mass, Balance, Volumetric equipment Measuring Seales / Tapes etc. are for scientific purpose only and should not be used for Trade / Commercial use.

CALIBRATED / CHECKED BY

(CALIBRATION ENGINEER)



AUTHORISED SIGNATORY

Tejb
TEJ BAHADUR