



LAB : Plot No. 272-C, Harbanspura, Near Govt. I.T.I., Yamuna Nagar - 135001, Haryana.  
Phone 01732-252120, 9991584585, 9050407120, E-mail : microtechnology.ind@gmail.com

## CALIBRATION CERTIFICATE

Customer Name And Address	M/S :Civil Hospital Ashta Central Processing Lab Bus Stand Sehore MP 466116	Certificate No.	MT/01/020923/MC2/00525
		ULR No.	CC333323000000524F
		SRF No. & Date	MT/01/02.09.23
		Receipt Date	02/09/2023
		Date of Calibration	02/09/2023
		Suggested Due Date	01/09/2024
		Date of Issue	02/09/2023

Instrument Details			
Instrument name	Micropipette	Sr. No.	YE209AS0035291
Make	Dragon	Location	Lab
Range / Size	5 – 50 µL	Accuracy	-----
Least Count	0.1 µL	Visual Inspection	OK
I.D. No.	CH ASHTA/CPL/DRAGON/03	Calibration Performed At	Lab

Detail of Reference Standard & Major Equipment Used			
Equipment Name	Digital Electronic Balance		
Make	Radwag		
Model / SR No.	AS 82/220. R2 / 640970		
Certificate No.	Tycon/WB/11/928		
Calibration Validity	16/11/2023		
Calibration By	Tycon Engineering		

Environmental Condition	Temperature	23± 1.5 °C	Calibration Reference	NABL129. ISO-8655-6
	Relative Humidity	50±10%	Calibration Procedure	CP/M/M&V/03

### Calibration Results


Serial No.	UUC in (µl)	Std. (Master) Value in (gm)	Std. (Master) Value Converted into (µl)	Uncertainty At 95% C.L. (coverage factor k=2)
01.	20	0.02008	20.08	±0.2µl
02.	30	0.03011	30.11	±0.2µl
03.	50	0.05016	50.16	±0.2µl

### Remarks:

<ul style="list-style-type: none"> <li>❖ (1) Standard equipment use for calibration are traceable to national/ international standards.</li> <li>❖ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor <math>k = 2</math> such that the coverage probability corresponds to approximately 95%.</li> <li>❖ (3) The above results are valid at the time of and under the stated conditions measurement.</li> <li>❖ (4) This certificate is refers only to the particular item submitted for calibration.</li> <li>❖ (5) Next calibration due date given as requested by the customer.</li> </ul>
--

Calibrated By  
(Calibration Engineer)  
(Anubha)

Format No. F01(7.8)



.....End of the Certificate.....

Approved By  
(Quality Manager)  
(Amit Saini)

Page No. 1 of 1