



TECHNICAL CALIBRATION REPORT

1. Name of Customer : **Diagnostica Span Pvt. Ltd. Nashik**
2. Name of instrument : **Erba H560**
3. Sr. No. of instrument : **K1104B2142117**
4. Date of Calibration : **27th Dec 2023**
5. Next Due Date of Calibration : **26th Dec 2024**

Voltage test point

Range	Observed
24 +/- 1.0 volt	24.00 volt
12 +/- 0.5 volt	12.11 volt
15 +/- 0.6 volt	14.50 volt
-15 +/- 0.6 volt	-15.02 volt
5 +/- 0.2 volt	5.2 volt

Pressure/ Vacuum Required	Pressure/ Vacuum Obtained
0.5Kg/cm2	0.52Kg/cm2
250 mmhg	250 mmhg

ALL ABOVE POINTS ARE WITH IN TECHNICAL SPECIFICATION.

For Hematology Parameters need H560 calibrator for calibration. .

For TRANSASIA BIO-MEDICALS LTD.

Pranil Kasar
Service Engineer

