



DIATEK

CALIBRATION CERTIFICATE

DIATEK HEALTHCARE PVT. LTD.
CIN: NO. U85100WB2010PTC142635
An ISO 9001:2015 Company

CUSTOMER NAME: **VELYS DIAGNOSTICS AND PATHOLOGY CENTER**

803, A SHRIDHAR ATHENS

ABOVE KALA MANDIR, SATTELITE

AHMEDABAD -

INSTRUMENT NAME	:	FULLY AUTOMATED BIOCHEMISTRY ANALYZER
MAKE	:	DIATEK
MODEL	:	Dia-CHEM-240
SERIAL NUMBER	:	CA24-201100440
DATE OF CALIBRATION	:	30/10/2023
NEXT DUE ON	:	29/04/2024

This is to certify that, we have calibrated FULLY AUTOMATED BIOCHEMISTRY ANALYZER,
Model: Dia-CHEM-240 made by DIATEK VELYS **DIAGNOSTICS AND PATHOLOGY CENTER**
on 30/10/2023.

Thanking you,

For, DIATEK HEALTHCARE PVT . LTD



CHIRAG SOLANKI,

SALES & SERVICE ENGINEER,

Contact No. +91 6351779284

Head Office : P-26, Kalindi Housing Scheme, Kolkata-700 089

Tele : 033-2522 3075 / 3448 / 4963. Fax : 033-2522-2401, E-mail : diatek@rediffmail.com

Website : www.diatek.in

Regd. Office / Works : P-54, Kalyani Industrial Estate, Phase-I, P.O/P.S : Kalyani, Dist : Nadia, W.B-741235



INSTALLATION QUALIFICATION (IQ)

NO	Description	Conformance to Requirement
Installation Qualification Preparation	Check the completeness of analyzer package and accessory box.	Yes
	Check that if logistics label status is normal or not.	Yes
	Unpack the instrument to check if the instrument appearance is normal or not.	Yes
	Instrument should be kept up-right, prevent slope.	Yes
	Open the accessory box, check each Accessory according to packing list.	Yes
Installation Condition Verification	Instrument installation place should keep away from interference source of strong electromagnetic.	Yes
	Instrument for indoor use only.	Yes
	Installation counter tops or ground should smooth and steady.	Yes
	Installation environment altitude below 2000 meters.	Yes
	Installation environment no corrosive gas and combustible gas.	Yes
	Instrument usage environment temperature: 10° C to 30° C, suggest to install air conditioner.	Yes
	Instrument assemble environment relative humidity: ≤70%.	Yes



Instrument assemble space required :	
The distance between back side of analyzer and wall	20cm
The distance between left & right side of analyzer and wall	Sufficient
The distance between front side of analyzer and other instrument	Sufficient
Keep enough space around power outlet position for convenient connection of power cable.	Sufficient
Instrument power requirement: Input voltage: 200-240V AC; (Frequency: 50Hz)	225 VAC
Voltage between Neutral and Ground	1.1 VAC 1.2
Analyzer and instrument should have UPS regulated power supply.	Connectd to UPS Power supply
Unwrap the instrument	Done
Connect system electrical circuit.	Done
Connect DI water tank	Done
Connect the waste liquid tank and tubings.	Done

Performed by: 

Date : 30/10/2023



OPERATIONAL QUALIFICATION (OQ)

NO.	Description	Conformance to Requirements
Operation Qualification Preparation	Switch on PC system.	Done
	Switch on Analyzer.	Done
	Check the connection between the PC and the instrument.	Done
	Fill the refrigeration unit with water as mentioned in the manual.	Done
	Program Glucose test. Place a blank reagent bottle filled with water. Place a sample cup filled with water. Program glucose test for 100 times. Run the test. (this helps to clear all the bubbles in the tubing's)	Performed.
	Program other parameters, calibrators and controls.	Performed

Performed by

Date: 30/10/2023



PERFORMANCE QUALIFICATION (PQ)

NO	Description	Conformance To requirements
Performance Qualification Preparation	<p>Initialize the machine Login the user software with correct user name and password ,and then machine initializes automatically.</p> <p>Wait 30 minutes after the initialization of the machine, to have the analyzer fully warm-up,</p> <p>Prepare the calibrator and control.</p> <p>Load the proper reagent bottles and at the reagent positions.</p> <p>Place the calibrator and control at the sample area. Program the parameters to be calibrated.</p> <p>Results are mentioned below :-</p> <p>Calibrated :</p> <p>Glucose = K Factor= 44.9740 Creatinine = K Factor = 150.2238 Cholesterol = K Factor = 22.1446</p> <p>Run control :</p> <p>Glucose : 269.8 (Tgt - 265.2; SD - 6.4) 76.6 (Tgt - 78.9; SD - 3.37) Creatinine : 5.02 (Tgt - 5.04; SD - 0.243) 1.99 (Tgt - 1.94; SD - 0.1) Cholesterol : 98.12 (Tgt - 97.57; SD - 2.89) 252.21 (Tgt - 246.05; SD - 5.72)</p>	Performed

Performed by: 

Date: 30/10/2023

Calibration Result(albumin)


Calibrator Name	Concentration	Unit	Average Response	Lot No.
WATER	0.00	g/dl	6.65	QFT10058
ELICAL-2	3.57	g/dl	5571.85	220782

Cal Method
Linear

Calibration Parameter
 $K = 1558.8775; b = 6.6532$

Reagent Blank
0.0000

Calibration time
10/30/2023 6:27:02 PM



Calibration Result(ALKP)


Calibrator Name	Concentration	Unit	Average Response	Lot No.
WATER	0.00	U/L	10.62	QFT10058
ELICAL-2	284.00	U/L	792.79	220782

Cal Method
Linear

Calibration Parameter
 $K = 2.9500; b = 10.6154$

Reagent Blank
0.0000

Calibration time
10/30/2023 6:33:03 PM



Calibration Result(Bil-Direct)


Calibrator Name	Concentration	Unit	Average Response	Lot No.
WATER	0.00	mg/dl	-114.63	QFT10058
ELICAL-2	2.27	mg/dl	1352.35	220782

Cal Method
Linear

Calibration Parameter
 $K = 646.2457; b = -114.6254$

Reagent Blank
0.0000

Calibration time
10/30/2023 6:30:47 PM



Calibration Result(CALCIUM)


Calibrator Name	Concentration	Unit	Average Response	Lot No.
WATER	0.00	mg/dl	-5.36	QFT10058
ELICAL-2	10.70	mg/dl	4078.98	220782

Cal Method
Linear

Calibration Parameter
 $K = 384.7100; b = -5.3589$

Reagent Blank
0.0000

Calibration time
10/30/2023 6:29:40 PM



Calibration Result(Bil-Total)


Calibrator Name	Concentration	Unit	Average Response	Lot No.
WATER	0.00	mg/dl	-62.19	QFT10058
ELICAL-2	4.15	mg/dl	634.99	220782

Cal Method
Linear

Calibration Parameter
 $K = 167.9938; b = -62.1853$

Reagent Blank
0.0000

Calibration time
10/30/2023 6:30:02 PM



Calibration Result(FBS)


Calibrator Name	Concentration	Unit	Average Response	Lot No.
WATER	0.00	mg/dl	-12.95	QFT10058
ELICAL-2	188.70	mg/dl	8473.64	220782

Cal Method
Linear

Calibration Parameter
 $K = 44.9740; b = -12.9498$

Reagent Blank
0.0000

Calibration time
10/30/2023 6:25:09 PM



Calibration Result(GGT)


Calibrator Name	Concentration	Unit	Average Response	Lot No.
WATER	0.00	U/L	19.78	QFT10058
ELICAL-2	114.00	U/L	377.78	220782

Cal Method
Linear

Calibration Parameter
 $K = 3.4200; b = 19.7824$

Reagent Blank
0.0000

Calibration time
10/30/2023 6:34:33 PM



Calibration Result(SGOT)


Calibrator Name	Concentration	Unit	Average Response	Lot No.
WATER	0.00	U/L	-8.37	QFT10058
ELICAL-2	110.40	U/L	-511.82	220782

Cal Method
Linear

Calibration Parameter
 $K = -4.5602; b = -8.3661$

Reagent Blank
0.0000

Calibration time
10/30/2023 6:28:32 PM



Calibration Result(SGPT)

Calibrator Name	Concentration	Unit	Average Response	Lot No.
WATER	0.00	U/L	-2.59	QFT10058
ELICAL-2	97.40	U/L	-498.59	220782

Cal Method

Linear

Calibration Parameter


$K = -5.0924; b = -2.5941$

Reagent Blank

0.0000

Calibration time

10/30/2023 6:27:47 PM



Calibration Result(TRIGLYCERIDE-DIATEK)


Calibrator Name	Concentration	Unit	Average Response	Lot No.
WATER	0.00	mg/dl	-10.75	QFT10058
ELICAL-2	146.00	mg/dl	2019.10	220782

Cal Method
Linear

Calibration Parameter
 $K = 13.9000; b = -10.7502$

Reagent Blank
0.0000

Calibration time
10/30/2023 6:24:16 PM



Calibration Result(CHOLESTEROL-DIATEK)


Calibrator Name	Concentration	Unit	Average Response	Lot No.
WATER	0.00	mg/dl	-0.71	QFT10058
ELICAL-2	177.00	mg/dl	3918.87	220782

Cal Method
Linear

Calibration Parameter
 $K = 22.1446; b = -0.7139$

Reagent Blank
0.0000

Calibration time
10/30/2023 6:23:26 PM



Calibration Result(UREA)


Calibrator Name	Concentration	Unit	Average Response	Lot No.
WATER	0.00	mg/dl	-36.82	QFT10058
ELICAL-2	100.00	mg/dl	-1249.77	220782

Cal Method
Linear

Calibration Parameter
 $K = -14,1300; b = -36.8200$

Reagent Blank
0.0000

Calibration time
10/30/2023 6:23:26 PM



Calibration Result(URIC ACID)

Calibrator Name	Concentration	Unit	Average Response	Lot No.
WATER	0.00	mg/dl	-28.68	QFT10058
ELICAL-2	5.39	mg/dl	987.01	220782


Cal Method
Linear

Calibration Parameter
 $K = 200.4400; b = -28.6807$

Reagent Blank
0.0000

Calibration time
10/30/2023 6:28:10 PM

Calibration Curve



Calibration Result(TOTAL Protein)


Calibrator Name	Concentration	Unit	Average Response	Lot No.
WATER	0.00	g/dl	0.89	QFT10058
ELICAL-2	5.26	g/dl	1648.72	220782

Cal Method
Linear

Calibration Parameter
 $K = 313.2762; b = 0.8856$

Reagent Blank
0.0000

Calibration time
10/30/2023 6:27:47 PM



Calibration Result(CRE-D)


Calibrator Name	Concentration	Unit	Average Response	Lot No.
WATER	0.00	mg/dl	5.98	QFT10058
ELICAL-2	4.03	mg/dl	651.16	220782

Cal Method
Linear

Calibration Parameter
 $K = 160.0928; b = 5.9839$

Reagent Blank
0.0000

Calibration time
10/30/2023 06:27:05 PM



Calibration Result(HDL)


Calibrator Name	Concentration	Unit	Average Response	Lot No.
WATER	0.00	mg/dl	-85.85	QFT10058
HDL CAL	66.28	mg/dl	2434.04	22-0829

Cal Method
Linear

Calibration Parameter
 $K = 38.0200; b = -85.8460$

Reagent Blank
0.0000

Calibration time
10/30/2023 07:15:43 PM



Calibration Result(HBA1C)

Calibrator Name	Concentration	Unit	Average Response	Lot No.
WATER	0.00	%	360.47	QFT10058
HBA1C 1	5.30	%	2017.00	
HBA1C 2	8.00	%	4989.00	
HBA1C 3	12.00	%	12318.51	
HBA1C 4	15.20	%	17133.47	

Cal Method
Spline

Calibration Parameter

$R0(1) = 360.4745; R0(2) = 2017.0000; R0(3) = 4989.0000; R0(4) = 12318.5139; a(1) = 312.5520; a(2) = 688.4438; a(3) = 1542.3506; a(4) = 1714.7711; b(1) = -70.9230; b(2) = 141.8460; b(3) = 174.4158; b(4) = -131.3107; c(1) = 13.3817; c(2) = 4.0210; c(3) = -25.4772; c(4) = 20.5173$

Reagent Blank
0.0000

Calibration time
10/30/2023 6:29:27 PM

