



CC-2725

**CERTIFICATE OF CALIBRATION**

Certificate No : CC-2324-S-1806/01

Page No : 1 of 2

Customer Name & Address  M/s. King Micro Lab  Ponnusami Complex, Opp.Autostand,  Chinnathirupathi Bus Stop, Salem - 8.	ULR No.	CC2725230000022347F
	SRF No.	VCI-2324-S-1806
	Date of Receipt	08.11.2023
	Date of Calibration	08.11.2023
	Due Date of Calibration	07.11.2024
	Date of Issue	09.11.2023

**Details of Unit Under Calibration**

Description	Centrifuge	Make/Model	Remi/C852
Range	0 to 3500 rpm, 10 to 60 Min.	Operating Range	500 to 3500 rpm, 10 to 60 Min.
Resolution	1 rpm, 10 Min.	Condition of UUC	Good
Serial Number	HHLC-11769	Instrument Location	--
ID Number	--	Cal. At	On Site

Environmental Condition		Calibration Method Used	
Temperature (°C)	23±5	National / International Standard	SANAS TR 45-02, NIST 960-12
Humidity (%RH)	40-70	Cal Procedure No	VCI-L03-WI-SPE-002, VCI-L03-WI-ET-005


**Standard Used**

Sl. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Digital Tachometer	VCI-M-TM-011	CC-23-24-L-002/08	Lutron / DT-2234C	Venus Calibration CC-2725	06.01.2024
2	Digital Timer	VCI-ET-DT-005	CC-23-24-L-002/02	M.TEC /CT7	Venus Calibration CC-2725	06.01.2024

**Remarks:**


- UUC is Defined as the Unit Under Calibration.
- This Calibration Certificate Results relates only to the above UUC & This report shall not be reproduced except in full/ part without approval of the Venus Calibration and Instruments.
- The reported results are valid at the time of under stated condition of measurement.
- Calibration of the UUC are traceable to National / International Standards.
- The recalibration interval shall be decided by the user.
- Nabl-133 guidelines are adopted for use of NABL symbol.
- The Measurement uncertainty is expressed at 95.45% confidence level with coverage factor  $k=2$ .

Checked &amp; Issued by:

  
R.Bhuvaneshwari  
(Engineer - Calibration)



Authorised by:

  
C.Dhinakaran  
(Technical / Quality Manager)