



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
All Analyser Summary Report - December 2023



Lab Name	LUPIN DIAGNOSITCS RRL HYDERABAD	Lab No	18894
Constituent Group	HbA1c	Date of Result Entered	18/12/2023
PT item	Lyophilized human whole blood based	Date of Report Published	02/01/2024

Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	HPLC-Ion Exchange	Biorad (D10 / D100 / Variant Turbo)	692	5.28	4.70	0.25	5.6 %	1.29	0.02
			Tosoh G8 - HPLC / HPLC-723G8)	209	5.20	5.29	0.28			0.04
			Any Analyser	175	5.07	9.65	0.49			0.07

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS:

LUPIN DIAGNOSITCS RRL HYDERABAD
1ST FLOOR, JAHANARA KAREEM COMPLEX, BALANAGAR
HYDERABAD
TELANGANA,500037

Pamela Christudoss

Coordinator Contact details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College,
Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****