

CALIBRATION CERTIFICATE

Page : 1 of 1

Certificate No : VCPL/CC/2024/05-12

Customer Name & Address : M/s SRI ABIRAMI CLINICAL LABORATORY, 331/38, TRICHY MAIN ROAD, DADAGAPATTI, SALEM-636 006.	SRF No	VCPL/24/05
	Date of Receipt	04-01-2024
	Date of Issue	05-01-2024
	Calibration Date	04-01-2024
	Next Due Date	03-01-2025
Customer Reference : Work Order	Calibration at	ON SITE

Details of Device Under Calibration

Description:	MICROSCOPE	Make:	LABOMED
Range:	-	Model/Type:	VISION 2000
Resolution:	-	Condition of DUC:	GOOD
Serial Number:	-	Location:	LABORATORY
ID number:	-	Cal. At:	ONSITE
Environmental Conditions		Calibration Method Used	
Temperature	25.9°C	National/International standard:	IEC60601-1
Humidity	51% Rh	Cal procedure no:	-

Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	Electrical Safety Analyzer	6014025	TSC/23-24/6682-1	FLUKE/ESA612	18-Jul-24


Electrical Safety Report

S.NO	PARAMETERS	MEASURED VALUES μ A	ALLOWABLE LIMITS μ A
1	Earth Resistance	0.196	<0.2 Ω
	Specification	MEASURED VALUES μ A	ALLOWABLE LIMITS μ a
2	Earth Leakage	196.2	<5000 μ A For B,BF,CF
	Specification	MEASURED VALUES μ A	ALLOWABLE LIMITS μ A
3	Enclosure Leakage	121.3	<500 μ A For B,BF,CF

Remarks:


- DUC is Defined as Device Under Calibration.
- Calibration Frequency decided by the customer.
- The above instrument is working in good condition.

Calibrated By:




M.KANNAN

(CALIBRATION ENGINEER)



Authorised by:



T. SUDHAKAR,

Technical Manager

Vanavil Calibrations Pvt Ltd.