

## CALIBRATION CERTIFICATE

Page : 1 of 1

Certificate No : VCPL/CC/2024/05-20

Customer Name & Address : M/s SRI ABIRAMI CLINICAL LABORATORY, 331/38, TRICHY MAIN ROAD, DADAGAPATTI, SALEM-636 006.	SRF No	VCPL/24/05
	Date of Receipt	04-01-2024
	Date of Issue	05-01-2024
	Calibration Date	04-01-2024
	Next Due Date	03-01-2025
Customer Reference : Work Order	Calibration at	ON SITE

### Details of Device Under Calibration

Description:	( SACL ) PROTEIN ANALYSER	Make:	AGAPPE
Range:	-	Model/Type:	MISPA- i2
Resolution:	-	Condition of DUC :	GOOD
Serial Number:	21171631823	Location:	LABORATORY
ID number:	-	Cal. At:	ONSITE
<b>Environmental Conditions</b>		<b>Calibration Method Used</b>	
Temperature	25.9°C	National/International standard:	IEC60601-1
Humidity	51% Rh	Cal procedure no:	-

### Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	Electrical Safety Analyzer	6014025	TSC/23-24/6682-1	FLUKE/ESA612	18-Jul-24


### Electrical Safety Report

S.NO	PARAMETERS	MEASURED VALUES $\mu$ A	ALLOWABLE LIMITS $\mu$ A
1	Earth Resistance	0.194	<0.2 $\Omega$
	<b>Specification</b>	<b>MEASURED VALUES <math>\mu</math>A</b>	<b>ALLOWABLE LIMITS <math>\mu</math>A</b>
2	Earth Leakage	196.4	<5000 $\mu$ A For B,BF,CF
	<b>Specification</b>	<b>MEASURED VALUES <math>\mu</math>A</b>	<b>ALLOWABLE LIMITS <math>\mu</math>A</b>
3	Enclosure Leakage	121.1	<500 $\mu$ A For B,BF,CF


### Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

  
M.KANNAN  
(CALIBRATION ENGINEER)

Authorised by:

  
T. SUDHAHAR,  
Technical Manager  
Vanavil Calibrations Pvt Ltd.

