



CALIBRATION REPORT

To,
Shree Vighnharta Superspeciality Hospital ,Dhule

DATE- 23/01/24

Dear Sir,

As desired by you please find enclosed herewith the Calibration report:

1. Instrument: Bio-chemistry Analyzer EC-5X S.N- **S2101152**.
2. Range of the instrument: Optical density Linear form 0-2.5.
3. Date of Calibration - 23 Jan 2024.
4. Next Calibration Due- 22 Jan 2025.

Quality Control Results-

Test Type	Test Results	Range
End Point (GLU)	99.01	80.2-108.2
Fix Time (URE)	33.3	29.85-37.85
Kinetic (SGPT)	44.2	37.5-52.5

All values are with in Range as per Quality control ERBA NORM.

Thanking you and assuring of our prompt attention at all times.

Yours faithfully,
For TRANSASIA BIO-MEDICALS LTD.

RAJESH WANKHEDE
REGIONAL SERVICE MANAGER

