



THECNICAL CALIBRATION REPORT

To,
Shree Vighnharta Superspeciality Hospital ,Dhule.

Date: 23 -Jan -2024

Dear Madam/Sir,

As desired by you please find herewith the Calibration report

1.Year of manufacturing : H360 , Sr.no. K10012047138

2.Date of Calibration : 23 -01- 2024

3.Next due date of calibration: 22-01-2025

Test Voltages / Test Sensors -

Test Voltage	Actual Voltage	Ok / Not Ok
Logic 5 V	5.2V	OK
Motor Voltage (23 – 25V)	24.1V	OK
Pressure (2.2 kg/cm ²)	2.21Kg/cm ²	OK
Vacuum 250Hg	250Hg	OK
Clog Voltage(100V)	102V	OK

Thanking you and assuring of our prompt attention at all times.

Yours faithfully,
For TRANSASIA BIO-MEDICALS LTD.

Rajesh Wankhede
Regional Service Manager

