



Since 1998

# NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058  
 Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370  
 E-mail : [ncqc@calibrationlaboratory.in](mailto:ncqc@calibrationlaboratory.in), • [calibrationlab.ncqc@gmail.com](mailto:calibrationlab.ncqc@gmail.com)

Visit our Web Site : [www.calibrationlaboratory.in](http://www.calibrationlaboratory.in)



CC-2128

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

## Calibration Certificate

Name of Customer → <b>Shree Vighnaharta Superspeciality Hospital</b> 1-22, SN 159, Nahata Nagar, Nakane Road Dhule - 424002		Certificate No. Date of Issue Date of Calibration Suggested Due Date	NCQC-MO/220523/02 23-05-2023 22-05-2023 21-05-2024
Date Of Receipt / Ref. No. → As Per Onsite Calibration		F/CR/M/070/1, Issue No.01 Page 1 of 1	
ULR – CC212823000002946F		Discipline → <b>Mechanical Calibration,</b> Acceleration & Speed	
Details of Observation of Unit Under Calibration		Identification No. : <b>SVH/CF-1</b> Serial No. : <b>=====</b> Name of Instrument : <b>RPM Of Centrifuge</b>	
Range	0 to 5000 RPM	Visual Inspection	OK
Least Count	Uneven	Initial error	Nil
Make / Model	Remi / C854/8	Location	Laboratory
Results of Calibration			
Reading Set on UUC in RPM	Reading Observed by Master Tachometer in RPM	Absolute Error in RPM	Expanded Uncertainty (±) in RPM
1000	1000.7	0.7	3.5
2000	2001.3	1.3	3.5
3000	3001.9	1.9	3.5
4000	4002.8	2.8	3.5
5000	5003.9	3.9	3.5
<b>Remarks:</b> ↳ Averages of minimum three readings are reported. ↳ Suggested due date is given based on customer requirements. ↳ Calibration points are given based on customer requirement. ↳ These results are obtained at the time of calibration. ↳ Location of Performance of Calibration → At site. ↳ Any hand written corrections (except @) or photocopies in the report invalidates this certificate. ↳ Average temperature → 25.5 °C and average humidity 52 % Rh during calibration of instruments. ↳ Environment condition during calibration: 25 ± 5°C, 40 to 60% Rh. ↳ Reference calibration method no.: NCQC/CM/M/070. ↳ Our masters are directly calibrated through NABL accredited calibration laboratory having direct traceability with national / international standard.			
Details of Master Instrument Used for Calibration			
Nomenclature	Make / Model	Id. No. / Sr. No.	Calibration Due Date
Master Digital Tachometer	Line Seiki / TM-7010	NCQC/M-154 / 184-0004	18-09-2023
NCQC System Certificate No.	Certificate No. & Traceability of Master with National Standards		
55	↳ Our master Digital Tacho Meter is calibrated and traceable to National Standard through NABL accredited Laboratory IDEMI Mumbai, Certificate no. CC/ECL/0909/22-23, Date – 19-09-2022.		

Traceable To National / International Standards.

Calibrated By Vishal Chauhan Reviewed & Approved By Jigar Panchal

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"