



NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058
Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370
E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com



CC-2128

Since 1998

Visit our Web Site : www.calibrationlaboratory.in

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate

Name of Customer → Shree Vighnaharta Superspeciality Hospital 1-22, SN 159, Nahata Nagar, Nakane Road Dhule - 424002		Certificate No. Date of Issue Date of Calibration Suggested Due Date	NCQC-MO/220523/03 23-05-2023 22-05-2023 21-05-2024
Date Of Receipt / Ref. No. → As Per Onsite Calibration		F/CR/M/070/1, Issue No.01 Page 1 of 1	
ULR – CC21282300002947F		Discipline → Mechanical Calibration, Acceleration & Speed	
Details of Observation of Unit Under Calibration		Identification No. : SVH/CF-2 Serial No. : ===== Name of Instrument : RPM Of Centrifuge	
Range	0 to 5000 RPM	Visual Inspection	OK
Least Count	Uneven	Initial error	Nil
Make / Model	Remi / R-303	Location	Laboratory
Results of Calibration			
Reading Set on UUC in RPM	Reading Observed by Master Tachometer In RPM	Absolute Error In RPM	Expanded Uncertainty (±) in RPM
1000	998.4	1.6	3.5
2000	1997.8	2.2	3.5
3000	2998.3	1.7	3.5
4000	3997.2	2.8	3.5
5000	4995.6	4.4	3.5
Remarks:			
<ul style="list-style-type: none"> ↘ Averages of minimum three readings are reported. ↘ Suggested due date is given based on customer requirements. ↘ Calibration points are given based on customer requirement. ↘ These results are obtained at the time of calibration. ↘ Location of Performance of Calibration → At site. ↘ Any hand written corrections (except @) or photocopies in the report invalidates this certificate. ↘ Average temperature → 25.7 °C and average humidity 50 % Rh during calibration of instruments. ↘ Environment condition during calibration: 25 ± 5°C, 40 to 60% Rh. ↘ Reference calibration method no.: NCQC/CM/M/070. ↘ Our masters are directly calibrated through NABL accredited calibration laboratory having direct traceability with national / international standard. 			
Details of Master Instrument Used for Calibration			
Nomenclature	Make / Model	Id. No. / Sr. No.	Calibration Due Date
Master Digital Tachometer	Line Seiki / TM-7010	NCQC/M-154 / 184-0004	18-09-2023
NCQC System Certificate No.	Certificate No. & Traceability of Master with National Standards		
55	↘ Our master Digital Tacho Meter is calibrated and traceable to National Standard through NABL accredited Laboratory IDEMI Mumbai, Certificate no. CC/ECL/0909/22-23, Date – 19-09-2022.		

Traceable To National / International Standards.

Calibrated By		Reviewed & Approved By	
	Vishal Chauhan		Jigar Panchal

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"