

CERTIFICATE OF CALIBRATION/ VERIFICATION

Customer Name : NRIUMSD-Hyderabad
Instrument : Haematology Analyzer
Model : Erba Elite 580
Serial Number : K11052310034
Calibration Date : 07-02-2024
Next Calibration due on before : 06-02-2025

This is to certify that the above-mentioned product has been verified of calibration for CBC 5 parameters (WBC, RBC, HGB, HCT and PLT) according to the standard procedures provided by Transasia Biomedical Ltd.

The reference instruments used for value-assignment are managed by the traceability system in Transasia Bio-Medical Ltd and these are traceable to the International Standards, such as ICSH.

Calibration/ Verification done by



Kalyan Haldar
Technical Service Department
Transasia Bio-Medicals Ltd.,
Sr.Application Manager
Hyderabad

Transasia Bio Medical Limited
No: 1-10-27/1, 1st Floor, Street No: 3, Ashok Nagar,
Hyderabad – 500020
Tel: 040-44309000/32949004 Email-sz4@transasia.co.in

An ISO 13485 & ISO 9001 Certified Co.

