



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - DECEMBER 2023



PC-1024

Lab Name **THE HEALTH COUNTY LABS** Lab No **18818**
Constituent Group **HbA1c** Date of Result Entered : **20/12/2023**
PT item **Lyophilized human whole blood based** Date of Report Published : **02/01/2024**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	HPLC-Ion Exchange	Any Analyser	175	5.07	9.65	0.49	4.6 %	-0.96	0.07

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

THE HEALTH COUNTY LABS

A UNIT OF HEALTH COUNTRY PVT LTD, D/50, UDYOG VIHAR PHASE - V, SEC -19

GURUGRAM

HARYANA122016

Pamela Christudoss

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****