

PIPETTE CALIBRATION REPORT

Product code: 4640090
 Description: Finnpiette F3 5-50µl
 Pipette SN: UW14388
 Tip: Finntip 250 Universal
 Nr. of channels: 1

Module SN: -

Manufacturer: Thermo Fisher Scientific, India Pvt. Ltd., D-96, C Road, MIDC Satpur, Nashik - 422007
 Tested by: 10222434

Date:	30/11/23	Time:	3:48:20 PM
Balance:	Mettler XP26PC	Balance SN:	B115129039, Calibrated 20.11.23
Environment device:	Almemo 2590-4AS	Env dev SN:	H19070544, Calibrated 07.10.23
Air temperature:	25.98 °C	Water temperature:	25.84 °C
Humidity:	69.3 %	Air pressure:	940.4 hPa
Z-factor:	1.0042		

Volume:	5 µl		50 µl
	1 4.94		1 50.10
	2 5.01		2 50.06
	3 5.01		3 50.08
	4 4.99		4 50.07
	5 4.96		5 50.07

Mean	4.98 µl	50.08 µl
A	-0.02 µl	0.08 µl
s	0.03 µl	0.01 µl
Acc%	-0.35	0.15
CV%	0.62	0.03
	PASSED	PASSED

Manufacturer specifications:

	Min. volume:	Max. volume:
	5 µl	50 µl
A±	0.15 µl	0.3 µl
s	0.125 µl	0.15 µl
Acc%±	3.00	0.60
CV%	2.50	0.30

Manufacturer specifications meet the requirements of ISO 8655