



In London, 7.5.2019

LAURA: Calibration procedure

To Whom It May Concern:

We Erba Lachema s.r.o. with registered office in Czech Republic, Karásek 2219/1D, 621 00 Brno, represented by its Product Manager Veronika Plešková hereby confirm as producer of LAURA instrument that this device is firmly calibrated by manufacturer and does not require any regular calibration by the user. Correct calibration of the instrument is controlled by grey strips provided initially together with the instrument.


Veronika Plešková



Product manager, Urine Analysis

Erba Corporate Service Ltd.
GW1 Great West House
Great West Road, London TW8 9DF



Erba Corporate Services Ltd . Registered in England and Wales . Company No: 11100097
Registered Office: c/o Erba Molecular, Unit 4 Bartholemews, Ely, Cambridgeshire, CB74EA, UK