



CALIBRATION CERTIFICATE

SP Well Fort
Trivandrum

Model Name & Sr. No : EM 200- 2002880
Range of the instrument: Optical density Linear form 0-2.5 A
Precision & Accuracy study on EM 200 done on 05.06.2023
Calibration Due Date: 04.06.2024
Calibration Done By : Mr. Jibin M T

Reagent/ Sample ratio -

Cup Position	Replicates	Absorbance (X)	SD	CV
Cup 1	20	1.5438 OD	0.0054	0.40%
Cup 2	20	1.5438 OD	0.0052	0.42 %
Cup 3	20	1.5449 OD	0.0047	0.38%

Test of Potassium Dichromate ($K_2Cr_2O_7$) on EM-200

Pipetted / Measured by Em200		Manually Pipettes Em200		Manually pipetted & read in Autospan	
Low	High	Low	High	Low	High
0.2005	1.8670	0.1975	1.8710	0.2025	1.7630





This instrument is manufactured in accordance with the European law in force, in particular to "Electricity Safety" EN 610101 and "Electromagnetic Compatibility" EN 61326 Standards.

	Ideal	Actual
Pressure	0.9 bar	0.90bar
Lamp Voltage	11.87 V	12.05V
Temperature	37 ⁰ C	37 ⁰ C

Photometer Calibration Check	Ok
Cuvette Blank Check	Ok

Wavelength	Off Set	Gain
340	255	540
405	255	525
505	255	520
546	255	487
578	255	535
600	255	470
660	255	485
700	255	475

The instrument does photometer check before each batch of tests is run. It will give an error message if the photometer test fails and it will not run any samples.

Thanking you and assuring of our prompt attention at all times.

Yours faithfully,
For Transasia Bio- Medicals Ltd,

For
Sajin
Sajin SNair
Regional Service Manager
Cochin



CERTIFICATE OF CALIBRATION

Customer Name	SP WELL FORT HOSPITAL	Certificate No.	SPW001
Customer Address	THIRUVANANTHAPURAM KERALA-695010	Date of Calibration	03/02/2024
		Valid Upto	02/02/2025
		Date of Issue	04/02/2024

DUC* Details			
DUC* Name	CENTRIFUGE	S No.	19491
Make/Model No.	ROTEK PLUS/P8E	Location	LABORATORY
ID No.	NA	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used		
Equipment Name	DIGITAL TACHO METER	THERMO HYGROMETER
Make	METRAVI	HTC-2
Model/SR No.	NCTM 1000/190518478	...
Equipment Id	CHPL/MT-KL/TAM/002	
Calibration certificate No	UCSL/07-23/202_05	UCSL/07-23/202_03
Calibration Validity	20/07/2024	20/07/2024
Calibration By	UNIQUE CALIBRATIONS	UNIQUE CALIBRATION SOLUTIONS

Environmental Condition			
Temperature	27°C	Relative Humidity	54%

Calibration Results				
SL NO	SET READING (RPM)	MEASURED READING (RPM)	TOLERANCE	STATUS
1	500	506	±5%	PASS
2	1000	1016	±5%	PASS
3	2500	2513	±5%	PASS

CALIBRATION STATUS	PASS	✓	FAIL	
---------------------------	-------------	---	-------------	--

REMARKS:- *DUC- DEVICE UNDER CALIBRATION

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By


ACHUTH P K



Issued and Approved By


PREJITH B
TECHNICAL MANAGER

CERTIFICATE OF CALIBRATION

Customer Name	SP WELL FORT HOSPITAL	Certificate No.	SPW001
Address	THIRUVANANTHAPURAM KERALA-695010	Date of Calibration	03/02/2024
		Valid Upto	02/02/2025
		Date of Issue	04/02/2024

DUC* Details					
DUC* Name	Make/Model No	S No.	ASSET ID	LOCATION	Visual Inspection
CENTRIFUGE	ROTEK PLUS/P8E	19491	NA	LABORATORY	OK

Details of Reference Standards & Major Equipments Used		
Equipment Name	ELECTRICAL SAFETY ANALYZER	THERMO HYGROMETER
Make	FLUKE	HTC-2
Model/SR No.	ESA 612/2959012	...
Equipment Id	CHPL/MT-KL/ESA/002	
Calibration certificate No	ET/240123C/01	UCSL/07-23/202_03
Calibration Validity	23/01/2025	20/07/2024
Calibration By	STIC	UNIQUE CALIBRATION SOLUTIONS

**STANDARD USED IS IEC 60601-1

ELECTRICAL SAFETY TEST VALUES

TEST NO.	NAME OF THE TEST	READING	HIGH LIMIT	LOW LIMIT	UNIT	STATUS
1	MAIN VOLTAGE -LIVE TO NEUTRAL	229	250	225	V	PASS
2	MAIN VOLTAGE -LIVE TO EARTH	228	240	220	V	PASS
3	MAIN VOLTAGE-NEUTRAL TO EARTH	0.3	3	0	V	PASS
4	EQUIPMENT CURRENT	0.4	16	0.1	mA	PASS
5	PROTECTIVE EARTH RESISTANCE	0.4	2	0	Ω	PASS
6	EARTH LEAKAGE CURRENT - NC	0.7	1	0.5	mA	PASS
7	EARTH LEAKAGE CURRENT - OC	0.6	1	0.5	mA	PASS
8	ENCLOSURE LEAKAGE CURRENT -NC	0.3	0.5	0.1	mA	PASS
9	ENCLOSURE LEAKAGE CURRENT -OC	0.2	0.5	0.1	mA	PASS
10	INSULATION RESITANCE					
	1.MAINS-PROTECTIVE EARTH	OL	100	0	MΩ	PASS
	2.APPLIED PART TO PROTECTIVE EARTH	NA	100	0	MΩ	PASS
	3.MAINS TO APPLIED PART	NA	100	0	MΩ	PASS
	4.MAINS TO NEUTRAL	OL	100	0	MΩ	PASS
	5.APPLIED PART TO NEUTRAL	NA	100	0	MΩ	PASS

*NOTE (NC -NORMAL CONDITION, OC-OPEN CONDTION), REMARKS- *DUC- DEVICE UNDER CALIBRATION

CALIBRATED BY

ACHUTH P



ISSUED AND APPROVED BY

PREJITH B
TECHNICAL MANAGER

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	SP WELLFORT HOSPITAL		
ADDRESS	ASHRMAM LN, SASTHAMANGALAM, THIRUVANANTHAPURAM , KERALA 695010	DATE OF CALIBRATION	08/02/2024
CERTIFICATE NO :	CHPL/PC/SPWH/001	VALID UP TO	07/02/2025
		DATE OF ISSUE	08/02/2024

INSTRUMENT DETAILS			
Instrument Name	MICRO PIPETTE	S.No.	---
Make / Model No.	THERMO SCIENTIFIC FINNPIPETTE F3	Visual Inspection	OK
LOCATION :	LAB	RANGE	100 µL
ASSET ID : NA			

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
Equipment Name	ANALYTICAL BALANCE	THERMO HYGROMETER
Make	WENSAR	HTC-1/-
Model / SR No.	MAB182/39736
Calibration Validity	27/07/2024	20/07/2024
Calibration By	UNIQUE	UNIQUE

ENVIRONMENTAL CONDITION			
Temperature	23.0 °C	Relative Humidity	45 %

Calibration Results							
SL NO	SET READING (µl)	MEASURED READING (µl)					STATUS
		I	II	III	IV	V	
1	100	100.00	100.01	100.02	99.99	99.98	PASS

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(CALIBRATION ENGINEER)

Approved By
(TECHNICAL MANAGER)
PREJITH T



CERTIFICATE OF CALIBRATION

CUSTOMER NAME	SP WELLFORT HOSPITAL		
ADDRESS	ASHRMAM LN, SASTHAMANGALAM, THIRUVANANTHAPURAM , KERALA 695010	DATE OF CALIBRATION	08/02/2024
CERTIFICATE NO :	CHPL/PC/SPWH/002	VALID UP TO	07/02/2025
		DATE OF ISSUE	08/02/2024

INSTRUMENT DETAILS			
Instrument Name	MICRO PIPETTE	S.No.	---
Make / Model No.	THERMO SCIENTIFIC FINNPIPETTE F3	Visual Inspection	OK
LOCATION :	LAB	RANGE	500 µL
ASSET ID : NA			

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
Equipment Name	ANALYTICAL BALANCE	THERMO HYGROMETER
Make	WENSAR	HTC-1/-
Model / SR No.	MAB182/39736
Calibration Validity	27/07/2024	20/07/2024
Calibration By	UNIQUE	UNIQUE

ENVIRONMENTAL CONDITION			
Temperature	23.0 °C	Relative Humidity	45 %

Calibration Results							
SL NO	SET READING (µl)	MEASURED READING (µl)					STATUS
		I	II	III	IV	V	
1	500	500.00	500.01	500.02	499.99	499.98	PASS

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(CALIBRATION ENGINEER)

Approved By
(TECHNICAL MANAGER)
PREJITH T B



CERTIFICATE OF CALIBRATION

CUSTOMER NAME	SP WELLFORT HOSPITAL		
ADDRESS	ASHRMAM LN, SASTHAMANGALAM, THIRUVANANTHAPURAM, KERALA 695010	DATE OF CALIBRATION	08/02/2024
CERTIFICATE NO :	CHPL/PC/SPWH/003	VALID UP TO	07/02/2025
		DATE OF ISSUE	08/02/2024

INSTRUMENT DETAILS			
Instrument Name	MICRO PIPETTE	S.No.	---
Make / Model No.	THERMO SCIENTIFIC FINNPIPETTE F3	Visual Inspection	OK
LOCATION :	LAB	RANGE	1000 µL
ASSET ID : NA			

DETAILS OF REFERENCE STANDARDS & MAJOR EQUIPMENTS USED		
Equipment Name	ANALYTICAL BALANCE	THERMO HYGROMETER
Make	WENSAR	HTC-1/-
Model / SR No.	MAB182/39736
Calibration Validity	27/07/2024	20/07/2024
Calibration By	UNIQUE	UNIQUE


ENVIRONMENTAL CONDITION			
Temperature	23.0 °C	Relative Humidity	45 %

Calibration Results							
SL NO	SET READING (µl)	MEASURED READING (µl)					STATUS
		I	II	III	IV	V	
1	1000	1000.00	1000.01	1000.02	999.99	999.98	PASS

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By 
(CALIBRATION ENGINEER)

Approved By
(TECHNICAL MANAGER)
PREJITH YB 





Certificate of Conformity and Calibration

TO WHOMSOEVER IT MAY CONCERN

ISO 15189:2012 REQUIREMENTS REGARDING, "CALIBRATION AND VERIFICATION PROCEDURE"

All Transasia Bio-Medicals Ltd diagnostics products which are distributed and for which a certificates is issued are CE Marked. Transasia Bio-Medicals Ltd , Mumbai ,manufacturers of diagnostic devices with company quality management system in compliance with standard ISO13485:2016 ;ISO9001:2015 .This means that all the processes in development and manufacturing of TBM products are guided by quality management system.

TBM declares and assure following

- The mentioned regulations require that production system and measuring devices are qualified and manufacturing and test procedures are validated as per ISO 13485:2016 and ISO 9001:2015 standard and it is assured through schedule maintenance and by regular qualification.
- TBM declares to have established procedure and to maintain it in order to assure the post marketing surveillance according to directive of 98/79/EC.
- All physical quantities, calibrators and controls used in TBM system are fully traceable to certified standards or reference materials.
- All TBM products are factory calibrated and final qc passed at the time of release.
- The performance of TBM system at customer site is assured if regular QC measurements, cleaning and maintenance procedure as described in the instruction for use or service documentation are performed
- Additional calibration or verification procedure is not required in order to assure the specified performances of every TBM system. Only if user deviates from the manufactures recommendation does he have to establish site specific calibration and verification procedure as part of his accreditation process.

Date:-26/04/2021

Manish Airan

Head Of Quality Department & Regulatory

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Feb-2024 To 09-Feb-2024

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
1	XYZ1234	BLANK	ALPU	0.0005 -		56097	6-Feb-2024 09:21:25
2	XYZ1234	BLANK	ALPU	0.0007 -		56098	6-Feb-2024 09:21:43
3	XYZ1234	BLANK	ALPU	0.0006 -		56099	6-Feb-2024 09:22:02
4	2301066B	MULTICAL	ALPU	0.0414 -		56100	6-Feb-2024 09:22:20
5	2301066B	MULTICAL	ALPU	0.0419 -		56101	6-Feb-2024 09:22:38
6	2301066B	MULTICAL	ALPU	0.0416 -		56102	6-Feb-2024 09:22:56
7	XYZ1234	BLANK	ALPU	0.0007 -		56123	6-Feb-2024 10:38:38
8	XYZ1234	BLANK	ALPU	0.0006 -		56124	6-Feb-2024 10:38:56
9	XYZ1234	BLANK	ALPU	0.0005 -		56125	6-Feb-2024 10:39:14
10	2301066B	MULTICAL	ALPU	0.0683 -		56126	6-Feb-2024 10:39:32
11	2301066B	MULTICAL	ALPU	0.0687 -		56127	6-Feb-2024 10:39:50
12	2301066B	MULTICAL	ALPU	0.0681 -		56128	6-Feb-2024 10:40:08

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Jan-2024 To 31-Jan-2024

Sr #	Lot #	Consumable	Test	Result	Unit	Flag	Curve #	Result Date
1	XYZ1234	BLANK	ALPU	0.0002	-		52783	3-Jan-2024 13:55:48
2	XYZ1234	BLANK	ALPU	0.0003	-		52784	3-Jan-2024 13:56:06
3	XYZ1234	BLANK	ALPU	0.0003	-		52785	3-Jan-2024 13:56:24
4	2301066B	MULTICAL	ALPU	0.0803	-		52786	3-Jan-2024 13:56:42
5	2301066B	MULTICAL	ALPU	NA	-	!R1	52787	3-Jan-2024 13:57:00
6	2301066B	MULTICAL	ALPU	0.0798	-		52788	3-Jan-2024 13:57:18
7	XYZ1234	BLANK	ALPU	0.0003	-		52842	4-Jan-2024 09:02:42
8	XYZ1234	BLANK	ALPU	0.0003	-		52843	4-Jan-2024 09:03:00
9	XYZ1234	BLANK	ALPU	0.0003	-		52844	4-Jan-2024 09:03:18
10	2301066B	MULTICAL	ALPU	0.0689	-		52845	4-Jan-2024 09:03:36
11	2301066B	MULTICAL	ALPU	0.0691	-		52846	4-Jan-2024 09:03:54
12	2301066B	MULTICAL	ALPU	0.0688	-		52847	4-Jan-2024 09:04:12
13	XYZ1234	BLANK	ALPU	0.0004	-		54886	26-Jan-2024 11:39:42
14	XYZ1234	BLANK	ALPU	0.0004	-		54887	26-Jan-2024 11:40:00
15	XYZ1234	BLANK	ALPU	0.0009	-		54888	26-Jan-2024 11:40:18
16	2301066B	MULTICAL	ALPU	0.0715	-		54889	26-Jan-2024 11:40:36
17	2301066B	MULTICAL	ALPU	0.0718	-		54890	26-Jan-2024 11:40:55

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Jan-2024 To 31-Jan-2024

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
18	2301066B	MULTICAL	ALPU	0.0715 -		54891	26-Jan-2024 11:41:13
19	XYZ1234	BLANK	UA	0.0029 -		54893	26-Jan-2024 11:41:48
20	XYZ1234	BLANK	UA	0.0030 -		54894	26-Jan-2024 11:42:06
21	XYZ1234	BLANK	UA	0.0030 -		54895	26-Jan-2024 11:42:24
22	2301066B	MULTICAL	UA	0.0721 -		54896	26-Jan-2024 11:42:42
23	2301066B	MULTICAL	UA	0.0721 -		54897	26-Jan-2024 11:43:00
24	2301066B	MULTICAL	UA	0.0721 -		54898	26-Jan-2024 11:43:19
25	XYZ1234	BLANK	BIT	0.0001 -		55140	29-Jan-2024 09:57:21
26	XYZ1234	BLANK	BIT	0.0001 -		55141	29-Jan-2024 09:57:39
27	XYZ1234	BLANK	BIT	0.0003 -		55142	29-Jan-2024 09:57:57
28	2301066B	MULTICAL	BIT	0.1768 -		55143	29-Jan-2024 09:58:15
29	2301066B	MULTICAL	BIT	0.1773 -		55144	29-Jan-2024 09:58:33
30	2301066B	MULTICAL	BIT	0.1770 -		55145	29-Jan-2024 09:58:51
31	XYZ1234	BLANK	UREA	-0.0038 -		55147	29-Jan-2024 09:59:27
32	XYZ1234	BLANK	UREA	-0.0042 -		55148	29-Jan-2024 09:59:45
33	XYZ1234	BLANK	UREA	-0.0037 -		55149	29-Jan-2024 10:00:03
34	2301066B	MULTICAL	UREA	-0.1425 -		55150	29-Jan-2024 10:00:21

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Jan-2024 To 31-Jan-2024

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
35	2301066B	MULTICAL	UREA	-0.1411 -		55151	29-Jan-2024 10:00:39
36	2301066B	MULTICAL	UREA	-0.1396 -		55152	29-Jan-2024 10:00:57
37	XYZ1234	BLANK	BIT	0.0002 -		55165	29-Jan-2024 12:54:07
38	XYZ1234	BLANK	BIT	0.0000 -		55166	29-Jan-2024 12:54:25
39	XYZ1234	BLANK	BIT	0.0007 -		55167	29-Jan-2024 12:54:44
40	2301066B	MULTICAL	BIT	0.1749 -		55168	29-Jan-2024 12:55:02
41	2301066B	MULTICAL	BIT	0.1766 -		55169	29-Jan-2024 12:55:20
42	2301066B	MULTICAL	BIT	0.1757 -		55170	29-Jan-2024 12:55:38

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Dec-2023 To 31-Jan-2024

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
1	XYZ1234	BLANK	UREA	-0.0065 -		49559	4-Dec-2023 05:41:33
2	XYZ1234	BLANK	UREA	-0.0063 -		49560	4-Dec-2023 05:41:51
3	XYZ1234	BLANK	UREA	-0.0066 -		49561	4-Dec-2023 05:42:10
4	2206133A	MULTICAL	UREA	-0.1593 -		49562	4-Dec-2023 05:42:28
5	2206133A	MULTICAL	UREA	-0.1619 -		49563	4-Dec-2023 05:42:46
6	2206133A	MULTICAL	UREA	-0.1623 -		49564	4-Dec-2023 05:43:04
7	XYZ1234	BLANK	UA	0.0024 -		49566	4-Dec-2023 05:43:40
8	XYZ1234	BLANK	UA	0.0029 -		49567	4-Dec-2023 05:43:58
9	XYZ1234	BLANK	UA	0.0030 -		49568	4-Dec-2023 05:44:16
10	2206133A	MULTICAL	UA	0.0836 -		49569	4-Dec-2023 05:44:34
11	2206133A	MULTICAL	UA	0.0831 -		49570	4-Dec-2023 05:44:52
12	2206133A	MULTICAL	UA	0.0834 -		49571	4-Dec-2023 05:45:10
13	XYZ1234	BLANK	UA	0.0022 -		49757	5-Dec-2023 04:52:44
14	XYZ1234	BLANK	UA	0.0026 -		49758	5-Dec-2023 04:53:02
15	XYZ1234	BLANK	UA	0.0029 -		49759	5-Dec-2023 04:53:20
16	2206133A	MULTICAL	UA	0.0824 -		49760	5-Dec-2023 04:53:38
17	2206133A	MULTICAL	UA	0.0827 -		49761	5-Dec-2023 04:53:56

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Dec-2023 To 31-Jan-2024

Sr #	Lot #	Consumable	Test	Result	Unit	Flag	Curve #	Result Date
18	2206133A	MULTICAL	UA	0.0824	-		49762	5-Dec-2023 04:54:15
19	XYZ1234	BLANK	CRPD	0.0005	-		49768	5-Dec-2023 11:16:53
20	XYZ1234	BLANK	CRPD	0.0013	-		49769	5-Dec-2023 11:17:12
21	XYZ1234	BLANK	CRPD	0.0011	-		49770	5-Dec-2023 11:17:30
22	S012337	CRPD CAL	CRPD	NA	-	S*	49771	5-Dec-2023 11:17:48
23	S012337	CRPD CAL	CRPD	NA	-	S*	49772	5-Dec-2023 11:18:06
24	S012337	CRPD CAL	CRPD	NA	-	S*	49773	5-Dec-2023 11:18:24
25	XYZ1234	BLANK	CRPD	0.0009	-		49789	5-Dec-2023 11:53:13
26	XYZ1234	BLANK	CRPD	0.0015	-		49790	5-Dec-2023 11:53:31
27	XYZ1234	BLANK	CRPD	0.0012	-		49791	5-Dec-2023 11:53:49
28	S012337	CRPD CAL	CRPD	0.9795	-		49792	5-Dec-2023 11:54:07
29	S012337	CRPD CAL	CRPD	0.9780	-		49793	5-Dec-2023 11:54:25
30	S012337	CRPD CAL	CRPD	0.9789	-		49794	5-Dec-2023 11:54:43
31	XYZ1234	BLANK	UA	0.0034	-		49853	6-Dec-2023 05:24:41
32	XYZ1234	BLANK	UA	0.0028	-		49854	6-Dec-2023 05:24:59
33	XYZ1234	BLANK	UA	0.0019	-		49855	6-Dec-2023 05:25:17
34	2206133A	MULTICAL	UA	0.0823	-		49856	6-Dec-2023 05:25:35

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Dec-2023 To 31-Jan-2024

Sr #	Lot #	Consumable	Test	Result	Unit	Flag	Curve #	Result Date
35	2206133A	MULTICAL	UA	0.0827	-		49857	6-Dec-2023 05:25:53
36	2206133A	MULTICAL	UA	0.0833	-		49858	6-Dec-2023 05:26:11
37	XYZ1234	BLANK	UA	0.0032	-		49969	7-Dec-2023 04:58:06
38	XYZ1234	BLANK	UA	0.0030	-		49970	7-Dec-2023 04:58:24
39	XYZ1234	BLANK	UA	0.0033	-		49971	7-Dec-2023 04:58:42
40	2206133A	MULTICAL	UA	0.0832	-		49972	7-Dec-2023 04:59:00
41	2206133A	MULTICAL	UA	0.0838	-		49973	7-Dec-2023 04:59:18
42	2206133A	MULTICAL	UA	0.0837	-		49974	7-Dec-2023 04:59:36
43	XYZ1234	BLANK	UREA	-0.0029	-		50518	12-Dec-2023 09:28:12
44	XYZ1234	BLANK	UREA	-0.0029	-		50519	12-Dec-2023 09:28:30
45	XYZ1234	BLANK	UREA	-0.0022	-		50520	12-Dec-2023 09:28:48
46	2206133A	MULTICAL	UREA	-0.1558	-		50521	12-Dec-2023 09:29:06
47	2206133A	MULTICAL	UREA	-0.1565	-		50522	12-Dec-2023 09:29:24
48	2206133A	MULTICAL	UREA	-0.1575	-		50523	12-Dec-2023 09:29:42
49	XYZ1234	BLANK	ALPU	0.0005	-	LV!	50903	15-Dec-2023 07:31:16
50	XYZ1234	BLANK	ALPU	0.0004	-		50904	15-Dec-2023 07:31:34
51	XYZ1234	BLANK	ALPU	0.0005	-		50905	15-Dec-2023 07:31:52

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type

: Calibration

01-Dec-2023 To 31-Jan-2024

Sr #	Lot #	Consumable	Test	Result	Unit	Flag	Curve #	Result Date
52	2206133A	MULTICAL	ALPU	0.0757	-		50906	15-Dec-2023 07:32:10
53	2206133A	MULTICAL	ALPU	0.0753	-		50907	15-Dec-2023 07:32:28
54	2206133A	MULTICAL	ALPU	0.0758	-		50908	15-Dec-2023 07:32:46
55	XYZ1234	BLANK	UREA	-0.0038	-		50910	15-Dec-2023 07:33:22
56	XYZ1234	BLANK	UREA	-0.0031	-		50911	15-Dec-2023 07:33:41
57	XYZ1234	BLANK	UREA	-0.0033	-		50912	15-Dec-2023 07:33:58
58	2206133A	MULTICAL	UREA	-0.1564	-		50913	15-Dec-2023 07:34:16
59	2206133A	MULTICAL	UREA	-0.1533	-		50914	15-Dec-2023 07:34:35
60	2206133A	MULTICAL	UREA	-0.1548	-		50915	15-Dec-2023 07:34:52
61	XYZ1234	BLANK	UA	0.0029	-		50917	15-Dec-2023 07:35:28
62	XYZ1234	BLANK	UA	0.0033	-		50918	15-Dec-2023 07:35:46
63	XYZ1234	BLANK	UA	0.0034	-		50919	15-Dec-2023 07:36:04
64	2206133A	MULTICAL	UA	0.0956	-		50920	15-Dec-2023 07:36:23
65	2206133A	MULTICAL	UA	0.0953	-		50921	15-Dec-2023 07:36:41
66	2206133A	MULTICAL	UA	0.0959	-		50922	15-Dec-2023 07:36:59
67	XYZ1234	BLANK	ALPU	0.0004	-		50924	15-Dec-2023 08:14:53
68	XYZ1234	BLANK	ALPU	0.0004	-		50925	15-Dec-2023 08:15:11

Print Date

09-Feb-2024

4 / 16

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Dec-2023 To 31-Jan-2024

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
69	XYZ1234	BLANK	ALPU	0.0005 -		50926	15-Dec-2023 08:15:29
70	2206133A	MULTICAL	ALPU	0.0770 -		50927	15-Dec-2023 08:15:48
71	2206133A	MULTICAL	ALPU	0.0772 -		50928	15-Dec-2023 08:16:06
72	2206133A	MULTICAL	ALPU	0.0771 -		50929	15-Dec-2023 08:16:24
73	XYZ1234	BLANK	UA	0.0031 -		50931	15-Dec-2023 08:17:00
74	XYZ1234	BLANK	UA	0.0035 -		50932	15-Dec-2023 08:17:18
75	XYZ1234	BLANK	UA	0.0033 -		50933	15-Dec-2023 08:17:36
76	2206133A	MULTICAL	UA	0.0965 -		50934	15-Dec-2023 08:17:54
77	2206133A	MULTICAL	UA	0.0946 -		50935	15-Dec-2023 08:18:12
78	2206133A	MULTICAL	UA	0.0945 -		50936	15-Dec-2023 08:18:30
79	XYZ1234	BLANK	UA	0.0035 -		50939	15-Dec-2023 09:00:13
80	XYZ1234	BLANK	UA	0.0034 -		50940	15-Dec-2023 09:00:31
81	XYZ1234	BLANK	UA	0.0033 -		50941	15-Dec-2023 09:00:49
82	2206133A	MULTICAL	UA	0.0948 -		50942	15-Dec-2023 09:01:07
83	2206133A	MULTICAL	UA	0.0959 -		50943	15-Dec-2023 09:01:25
84	2206133A	MULTICAL	UA	0.0951 -		50944	15-Dec-2023 09:01:43
85	XYZ1234	BLANK	UA	0.0036 -		50962	15-Dec-2023 11:53:49

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type

: Calibration

01-Dec-2023 To 31-Jan-2024

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
86	XYZ1234	BLANK	UA	0.0035 -		50963	15-Dec-2023 11:54:07
87	XYZ1234	BLANK	UA	0.0035 -		50964	15-Dec-2023 11:54:25
88	2206133A	MULTICAL	UA	0.0813 -		50965	15-Dec-2023 11:54:43
89	2206133A	MULTICAL	UA	0.0817 -		50966	15-Dec-2023 11:55:01
90	2206133A	MULTICAL	UA	0.0824 -		50967	15-Dec-2023 11:55:19
91	XYZ1234	BLANK	SGPTD	-0.0008 -		51230	18-Dec-2023 11:58:26
92	XYZ1234	BLANK	SGPTD	-0.0007 -		51231	18-Dec-2023 11:58:44
93	XYZ1234	BLANK	SGPTD	-0.0006 -		51232	18-Dec-2023 11:59:02
94	2206133A	MULTICAL	SGPTD	-0.0348 -		51233	18-Dec-2023 11:59:20
95	2206133A	MULTICAL	SGPTD	-0.0347 -		51234	18-Dec-2023 11:59:38
96	2206133A	MULTICAL	SGPTD	-0.0349 -		51235	18-Dec-2023 11:59:57
97	XYZ1234	BLANK	BIDD	0.0000 -		51237	18-Dec-2023 12:00:33
98	XYZ1234	BLANK	BIDD	-0.0002 -		51238	18-Dec-2023 12:00:51
99	XYZ1234	BLANK	BIDD	0.0002 -		51239	18-Dec-2023 12:01:09
100	2206133A	MULTICAL	BIDD	0.1044 -		51240	18-Dec-2023 12:01:27
101	2206133A	MULTICAL	BIDD	0.1042 -		51241	18-Dec-2023 12:01:45
102	2206133A	MULTICAL	BIDD	0.1043 -		51242	18-Dec-2023 12:02:03

Print Date

09-Feb-2024

6 / 16

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Dec-2023 To 31-Jan-2024

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
103	XYZ1234	BLANK	UA	0.0025 -		51244	18-Dec-2023 12:02:39
104	XYZ1234	BLANK	UA	0.0025 -		51245	18-Dec-2023 12:02:57
105	XYZ1234	BLANK	UA	0.0030 -		51246	18-Dec-2023 12:03:15
106	2206133A	MULTICAL	UA	0.0815 -		51247	18-Dec-2023 12:03:33
107	2206133A	MULTICAL	UA	0.0809 -		51248	18-Dec-2023 12:03:51
108	2206133A	MULTICAL	UA	0.0813 -		51249	18-Dec-2023 12:04:09
109	XYZ1234	BLANK	SGPTD	-0.0006 -		51255	18-Dec-2023 12:33:35
110	XYZ1234	BLANK	SGPTD	-0.0004 -		51256	18-Dec-2023 12:33:53
111	XYZ1234	BLANK	SGPTD	-0.0003 -		51257	18-Dec-2023 12:34:11
112	2206133A	MULTICAL	SGPTD	-0.0346 -		51258	18-Dec-2023 12:34:29
113	2206133A	MULTICAL	SGPTD	-0.0343 -		51259	18-Dec-2023 12:34:47
114	2206133A	MULTICAL	SGPTD	-0.0351 -		51260	18-Dec-2023 12:35:05
115	XYZ1234	BLANK	BIDD	0.0003 -		51262	18-Dec-2023 12:35:41
116	XYZ1234	BLANK	BIDD	0.0003 -		51263	18-Dec-2023 12:35:59
117	XYZ1234	BLANK	BIDD	0.0004 -		51264	18-Dec-2023 12:36:18
118	2206133A	MULTICAL	BIDD	0.1051 -		51265	18-Dec-2023 12:36:36
119	2206133A	MULTICAL	BIDD	0.1047 -		51266	18-Dec-2023 12:36:54

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Dec-2023 To 31-Jan-2024

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
120	2206133A	MULTICAL	BIDD	0.1048 -		51267	18-Dec-2023 12:37:12
121	XYZ1234	BLANK	UA	0.0033 -		51269	18-Dec-2023 12:37:48
122	XYZ1234	BLANK	UA	0.0036 -		51270	18-Dec-2023 12:38:06
123	XYZ1234	BLANK	UA	0.0038 -		51271	18-Dec-2023 12:38:24
124	2206133A	MULTICAL	UA	0.0823 -		51272	18-Dec-2023 12:38:42
125	2206133A	MULTICAL	UA	0.0827 -		51273	18-Dec-2023 12:39:00
126	2206133A	MULTICAL	UA	0.0819 -		51274	18-Dec-2023 12:39:18
127	XYZ1234	BLANK	SGPTD	-0.0005 -		51276	18-Dec-2023 13:03:33
128	XYZ1234	BLANK	SGPTD	-0.0001 -		51277	18-Dec-2023 13:03:51
129	XYZ1234	BLANK	SGPTD	-0.0004 -		51278	18-Dec-2023 13:04:09
130	2206133A	MULTICAL	SGPTD	-0.0348 -		51279	18-Dec-2023 13:04:28
131	2206133A	MULTICAL	SGPTD	-0.0343 -		51280	18-Dec-2023 13:04:46
132	2206133A	MULTICAL	SGPTD	-0.0348 -		51281	18-Dec-2023 13:05:04
133	XYZ1234	BLANK	UA	0.0036 -		51283	18-Dec-2023 13:05:40
134	XYZ1234	BLANK	UA	0.0034 -		51284	18-Dec-2023 13:05:58
135	XYZ1234	BLANK	UA	0.0080 -		51285	18-Dec-2023 13:06:16
136	2206133A	MULTICAL	UA	0.0824 -		51286	18-Dec-2023 13:06:34

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Dec-2023 To 31-Jan-2024

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
137	2206133A	MULTICAL	UA	0.0826 -		51287	18-Dec-2023 13:06:52
138	2206133A	MULTICAL	UA	0.0824 -		51288	18-Dec-2023 13:07:10
139	XYZ1234	BLANK	PROT	0.0825 -		51520	20-Dec-2023 11:34:33
140	XYZ1234	BLANK	PROT	0.0806 -		51521	20-Dec-2023 11:34:51
141	XYZ1234	BLANK	PROT	0.0818 -		51522	20-Dec-2023 11:35:09
142	2206133A	MULTICAL	PROT	0.4251 -		51523	20-Dec-2023 11:35:27
143	2206133A	MULTICAL	PROT	0.4252 -		51524	20-Dec-2023 11:35:45
144	2206133A	MULTICAL	PROT	0.4246 -		51525	20-Dec-2023 11:36:03
145	XYZ1234	BLANK	ALPU	0.0004 -		51916	23-Dec-2023 04:42:56
146	XYZ1234	BLANK	ALPU	0.0004 -		51917	23-Dec-2023 04:43:14
147	XYZ1234	BLANK	ALPU	0.0004 -		51918	23-Dec-2023 04:43:32
148	2206133A	MULTICAL	ALPU	0.0736 -		51919	23-Dec-2023 04:43:50
149	2206133A	MULTICAL	ALPU	0.0737 -		51920	23-Dec-2023 04:44:08
150	2206133A	MULTICAL	ALPU	0.0733 -		51921	23-Dec-2023 04:44:27
151	XYZ1234	BLANK	UREA	-0.0045 -		51923	23-Dec-2023 04:45:03
152	XYZ1234	BLANK	UREA	-0.0033 -		51924	23-Dec-2023 04:45:21
153	XYZ1234	BLANK	UREA	-0.0034 -		51925	23-Dec-2023 04:45:39

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Dec-2023 To 31-Jan-2024

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
154	2206133A	MULTICAL	UREA	-0.1493 -		51926	23-Dec-2023 04:45:57
155	2206133A	MULTICAL	UREA	-0.1465 -		51927	23-Dec-2023 04:46:15
156	2206133A	MULTICAL	UREA	-0.1515 -		51928	23-Dec-2023 04:46:33
157	XYZ1234	BLANK	UA	0.0028 -		51930	23-Dec-2023 04:47:09
158	XYZ1234	BLANK	UA	0.0029 -		51931	23-Dec-2023 04:47:27
159	XYZ1234	BLANK	UA	0.0034 -		51932	23-Dec-2023 04:47:45
160	2206133A	MULTICAL	UA	0.0819 -		51933	23-Dec-2023 04:48:03
161	2206133A	MULTICAL	UA	0.0821 -		51934	23-Dec-2023 04:48:21
162	2206133A	MULTICAL	UA	0.0807 -		51935	23-Dec-2023 04:48:39
163	XYZ1234	BLANK	UA	0.0030 -		51997	24-Dec-2023 04:42:01
164	XYZ1234	BLANK	UA	0.0022 -		51998	24-Dec-2023 04:42:19
165	XYZ1234	BLANK	UA	0.0031 -		51999	24-Dec-2023 04:42:37
166	2206133A	MULTICAL	UA	0.0826 -		52000	24-Dec-2023 04:42:55
167	2206133A	MULTICAL	UA	0.0833 -		52001	24-Dec-2023 04:43:13
168	2206133A	MULTICAL	UA	0.0824 -		52002	24-Dec-2023 04:43:32
169	XYZ1234	BLANK	ALPU	0.0004 -		52054	25-Dec-2023 05:09:15
170	XYZ1234	BLANK	ALPU	0.0003 -		52055	25-Dec-2023 05:09:33

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Dec-2023 To 31-Jan-2024

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
171	XYZ1234	BLANK	ALPU	0.0000 -		52056	25-Dec-2023 05:09:51
172	2206133A	MULTICAL	ALPU	0.0717 -		52057	25-Dec-2023 05:10:09
173	2206133A	MULTICAL	ALPU	0.0715 -		52058	25-Dec-2023 05:10:27
174	2206133A	MULTICAL	ALPU	0.0708 -		52059	25-Dec-2023 05:10:45
175	XYZ1234	BLANK	UA	0.0028 -		52061	25-Dec-2023 05:11:21
176	XYZ1234	BLANK	UA	0.0032 -		52062	25-Dec-2023 05:11:39
177	XYZ1234	BLANK	UA	0.0032 -		52063	25-Dec-2023 05:11:57
178	2206133A	MULTICAL	UA	0.0832 -		52064	25-Dec-2023 05:12:15
179	2206133A	MULTICAL	UA	0.0822 -		52065	25-Dec-2023 05:12:33
180	2206133A	MULTICAL	UA	0.0831 -		52066	25-Dec-2023 05:12:51
181	XYZ1234	BLANK	ALPU	0.0003 -		52121	26-Dec-2023 05:29:16
182	XYZ1234	BLANK	ALPU	0.0003 -		52122	26-Dec-2023 05:29:34
183	XYZ1234	BLANK	ALPU	0.0003 -		52123	26-Dec-2023 05:29:52
184	2206133A	MULTICAL	ALPU	0.0698 -		52124	26-Dec-2023 05:30:10
185	2206133A	MULTICAL	ALPU	0.0698 -		52125	26-Dec-2023 05:30:28
186	2206133A	MULTICAL	ALPU	0.0696 -		52126	26-Dec-2023 05:30:46
187	XYZ1234	BLANK	UA	0.0032 -		52128	26-Dec-2023 05:31:22

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Dec-2023 To 31-Jan-2024

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
188	XYZ1234	BLANK	UA	0.0050 -		52129	26-Dec-2023 05:31:40
189	XYZ1234	BLANK	UA	0.0032 -		52130	26-Dec-2023 05:31:58
190	2206133A	MULTICAL	UA	0.0829 -		52131	26-Dec-2023 05:32:16
191	2206133A	MULTICAL	UA	0.0834 -		52132	26-Dec-2023 05:32:34
192	2206133A	MULTICAL	UA	0.0831 -		52133	26-Dec-2023 05:32:52
193	XYZ1234	BLANK	UA	0.0032 -		52135	26-Dec-2023 07:05:00
194	XYZ1234	BLANK	UA	0.0032 -		52136	26-Dec-2023 07:05:18
195	XYZ1234	BLANK	UA	0.0029 -		52137	26-Dec-2023 07:05:36
196	2206133A	MULTICAL	UA	0.0839 -		52138	26-Dec-2023 07:05:54
197	2206133A	MULTICAL	UA	0.0843 -		52139	26-Dec-2023 07:06:13
198	2206133A	MULTICAL	UA	0.0841 -		52140	26-Dec-2023 07:06:30
199	XYZ1234	BLANK	UA	0.0029 -		52158	26-Dec-2023 12:34:56
200	XYZ1234	BLANK	UA	0.0027 -		52159	26-Dec-2023 12:35:14
201	XYZ1234	BLANK	UA	0.0028 -		52160	26-Dec-2023 12:35:32
202	2206133A	MULTICAL	UA	0.0861 -		52161	26-Dec-2023 12:35:50
203	2206133A	MULTICAL	UA	0.0868 -		52162	26-Dec-2023 12:36:08
204	2206133A	MULTICAL	UA	0.0867 -		52163	26-Dec-2023 12:36:27

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Dec-2023 To 31-Jan-2024

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
205	XYZ1234	BLANK	UA	0.0036 -		52187	27-Dec-2023 04:56:48
206	XYZ1234	BLANK	UA	0.0031 -		52188	27-Dec-2023 04:57:06
207	XYZ1234	BLANK	UA	0.0030 -		52189	27-Dec-2023 04:57:24
208	2206133A	MULTICAL	UA	0.0859 -		52190	27-Dec-2023 04:57:42
209	2206133A	MULTICAL	UA	0.0855 -		52191	27-Dec-2023 04:58:00
210	2206133A	MULTICAL	UA	0.0858 -		52192	27-Dec-2023 04:58:18
211	XYZ1234	BLANK	UA	0.0029 -		52281	28-Dec-2023 05:04:21
212	XYZ1234	BLANK	UA	0.0033 -		52282	28-Dec-2023 05:04:39
213	XYZ1234	BLANK	UA	0.0029 -		52283	28-Dec-2023 05:04:58
214	2206133A	MULTICAL	UA	0.0885 -		52284	28-Dec-2023 05:05:15
215	2206133A	MULTICAL	UA	0.0885 -		52285	28-Dec-2023 05:05:33
216	2206133A	MULTICAL	UA	0.0889 -		52286	28-Dec-2023 05:05:52
217	XYZ1234	BLANK	UA	0.0045 -		52477	30-Dec-2023 11:56:13
218	XYZ1234	BLANK	UA	0.0039 -		52478	30-Dec-2023 11:56:31
219	XYZ1234	BLANK	UA	0.0041 -		52479	30-Dec-2023 11:56:49
220	2206133A	MULTICAL	UA	0.0839 -		52480	30-Dec-2023 11:57:07
221	2206133A	MULTICAL	UA	0.0845 -		52481	30-Dec-2023 11:57:25

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Dec-2023 To 31-Jan-2024

Sr #	Lot #	Consumable	Test	Result	Unit	Flag	Curve #	Result Date
222	2206133A	MULTICAL	UA	0.0845	-		52482	30-Dec-2023 11:57:43
223	XYZ1234	BLANK	ALPU	0.0002	-		52783	3-Jan-2024 13:55:48
224	XYZ1234	BLANK	ALPU	0.0003	-		52784	3-Jan-2024 13:56:06
225	XYZ1234	BLANK	ALPU	0.0003	-		52785	3-Jan-2024 13:56:24
226	2301066B	MULTICAL	ALPU	0.0803	-		52786	3-Jan-2024 13:56:42
227	2301066B	MULTICAL	ALPU	NA	-	!R1	52787	3-Jan-2024 13:57:00
228	2301066B	MULTICAL	ALPU	0.0798	-		52788	3-Jan-2024 13:57:18
229	XYZ1234	BLANK	ALPU	0.0003	-		52842	4-Jan-2024 09:02:42
230	XYZ1234	BLANK	ALPU	0.0003	-		52843	4-Jan-2024 09:03:00
231	XYZ1234	BLANK	ALPU	0.0003	-		52844	4-Jan-2024 09:03:18
232	2301066B	MULTICAL	ALPU	0.0689	-		52845	4-Jan-2024 09:03:36
233	2301066B	MULTICAL	ALPU	0.0691	-		52846	4-Jan-2024 09:03:54
234	2301066B	MULTICAL	ALPU	0.0688	-		52847	4-Jan-2024 09:04:12
235	XYZ1234	BLANK	ALPU	0.0004	-		54886	26-Jan-2024 11:39:42
236	XYZ1234	BLANK	ALPU	0.0004	-		54887	26-Jan-2024 11:40:00
237	XYZ1234	BLANK	ALPU	0.0009	-		54888	26-Jan-2024 11:40:18
238	2301066B	MULTICAL	ALPU	0.0715	-		54889	26-Jan-2024 11:40:36

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Dec-2023 To 31-Jan-2024

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
239	2301066B	MULTICAL	ALPU	0.0718 -		54890	26-Jan-2024 11:40:55
240	2301066B	MULTICAL	ALPU	0.0715 -		54891	26-Jan-2024 11:41:13
241	XYZ1234	BLANK	UA	0.0029 -		54893	26-Jan-2024 11:41:48
242	XYZ1234	BLANK	UA	0.0030 -		54894	26-Jan-2024 11:42:06
243	XYZ1234	BLANK	UA	0.0030 -		54895	26-Jan-2024 11:42:24
244	2301066B	MULTICAL	UA	0.0721 -		54896	26-Jan-2024 11:42:42
245	2301066B	MULTICAL	UA	0.0721 -		54897	26-Jan-2024 11:43:00
246	2301066B	MULTICAL	UA	0.0721 -		54898	26-Jan-2024 11:43:19
247	XYZ1234	BLANK	BIT	0.0001 -		55140	29-Jan-2024 09:57:21
248	XYZ1234	BLANK	BIT	0.0001 -		55141	29-Jan-2024 09:57:39
249	XYZ1234	BLANK	BIT	0.0003 -		55142	29-Jan-2024 09:57:57
250	2301066B	MULTICAL	BIT	0.1768 -		55143	29-Jan-2024 09:58:15
251	2301066B	MULTICAL	BIT	0.1773 -		55144	29-Jan-2024 09:58:33
252	2301066B	MULTICAL	BIT	0.1770 -		55145	29-Jan-2024 09:58:51
253	XYZ1234	BLANK	UREA	-0.0038 -		55147	29-Jan-2024 09:59:27
254	XYZ1234	BLANK	UREA	-0.0042 -		55148	29-Jan-2024 09:59:45
255	XYZ1234	BLANK	UREA	-0.0037 -		55149	29-Jan-2024 10:00:03

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type

: Calibration

01-Dec-2023 To 31-Jan-2024

Sr #	Lot #	Consumable	Test	Result	Unit	Flag	Curve #	Result Date
256	2301066B	MULTICAL	UREA	-0.1425	-		55150	29-Jan-2024 10:00:21
257	2301066B	MULTICAL	UREA	-0.1411	-		55151	29-Jan-2024 10:00:39
258	2301066B	MULTICAL	UREA	-0.1396	-		55152	29-Jan-2024 10:00:57
259	XYZ1234	BLANK	BIT	0.0002	-		55165	29-Jan-2024 12:54:07
260	XYZ1234	BLANK	BIT	0.0000	-		55166	29-Jan-2024 12:54:25
261	XYZ1234	BLANK	BIT	0.0007	-		55167	29-Jan-2024 12:54:44
262	2301066B	MULTICAL	BIT	0.1749	-		55168	29-Jan-2024 12:55:02
263	2301066B	MULTICAL	BIT	0.1766	-		55169	29-Jan-2024 12:55:20
264	2301066B	MULTICAL	BIT	0.1757	-		55170	29-Jan-2024 12:55:38

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Nov-2023 To 30-Nov-2023

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
1	XYZ1234	BLANK	UA	0.0034 -		46163	2-Nov-2023 05:47:50
2	XYZ1234	BLANK	UA	0.0032 -		46164	2-Nov-2023 05:48:08
3	XYZ1234	BLANK	UA	0.0032 -		46165	2-Nov-2023 05:48:27
4	2206133A	MULTICAL	UA	0.0921 -		46166	2-Nov-2023 05:48:45
5	2206133A	MULTICAL	UA	0.0920 -		46167	2-Nov-2023 05:49:03
6	2206133A	MULTICAL	UA	0.0921 -		46168	2-Nov-2023 05:49:21
7	XYZ1234	BLANK	UA	0.0035 -		46170	2-Nov-2023 06:10:54
8	XYZ1234	BLANK	UA	0.0031 -		46171	2-Nov-2023 06:11:11
9	XYZ1234	BLANK	UA	0.0029 -		46172	2-Nov-2023 06:11:29
10	2206133A	MULTICAL	UA	0.0914 -		46173	2-Nov-2023 06:11:47
11	2206133A	MULTICAL	UA	0.0919 -		46174	2-Nov-2023 06:12:06
12	2206133A	MULTICAL	UA	0.0918 -		46175	2-Nov-2023 06:12:24
13	XYZ1234	BLANK	UA	0.0029 -		46178	2-Nov-2023 08:31:33
14	XYZ1234	BLANK	UA	0.0045 -		46179	2-Nov-2023 08:31:51
15	XYZ1234	BLANK	UA	0.0030 -		46180	2-Nov-2023 08:32:09
16	2206133A	MULTICAL	UA	0.0928 -		46181	2-Nov-2023 08:32:27
17	2206133A	MULTICAL	UA	0.0931 -		46182	2-Nov-2023 08:32:45

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Nov-2023 To 30-Nov-2023

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
18	2206133A	MULTICAL	UA	0.0929 -		46183	2-Nov-2023 08:33:03
19	XYZ1234	BLANK	UA	0.0032 -		46305	2-Nov-2023 15:06:19
20	XYZ1234	BLANK	UA	0.0033 -		46306	2-Nov-2023 15:06:37
21	XYZ1234	BLANK	UA	0.0033 -		46307	2-Nov-2023 15:06:55
22	2206133A	MULTICAL	UA	0.0945 -		46308	2-Nov-2023 15:07:13
23	2206133A	MULTICAL	UA	0.0954 -		46309	2-Nov-2023 15:07:31
24	2206133A	MULTICAL	UA	0.0947 -		46310	2-Nov-2023 15:07:49
25	XYZ1234	BLANK	UA	0.0058 -		46353	3-Nov-2023 09:42:36
26	XYZ1234	BLANK	UA	0.0030 -		46354	3-Nov-2023 09:42:54
27	XYZ1234	BLANK	UA	0.0031 -		46355	3-Nov-2023 09:43:12
28	2206133A	MULTICAL	UA	0.1051 -		46356	3-Nov-2023 09:43:30
29	2206133A	MULTICAL	UA	0.1064 -		46357	3-Nov-2023 09:43:48
30	2206133A	MULTICAL	UA	0.1063 -		46358	3-Nov-2023 09:44:06
31	XYZ1234	BLANK	LDL	0.0015 -		46374	3-Nov-2023 14:10:46
32	XYZ1234	BLANK	LDL	0.0012 -		46375	3-Nov-2023 14:11:04
33	XYZ1234	BLANK	LDL	0.0010 -		46376	3-Nov-2023 14:11:22
34	S072329	HDL LDL CAL	LDL	0.0893 -		46377	3-Nov-2023 14:11:40

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Nov-2023 To 30-Nov-2023

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
35	S072329	HDL LDL CAL	LDL	0.0896 -		46378	3-Nov-2023 14:11:58
36	S072329	HDL LDL CAL	LDL	0.0872 -		46379	3-Nov-2023 14:12:17
37	XYZ1234	BLANK	ALPU	0.0004 -		47153	10-Nov-2023 04:55:24
38	XYZ1234	BLANK	ALPU	0.0004 -		47154	10-Nov-2023 04:55:42
39	XYZ1234	BLANK	ALPU	0.0004 -		47155	10-Nov-2023 04:56:00
40	2206133A	MULTICAL	ALPU	0.1014 -		47156	10-Nov-2023 04:56:18
41	2206133A	MULTICAL	ALPU	0.1017 -		47157	10-Nov-2023 04:56:36
42	2206133A	MULTICAL	ALPU	0.1013 -		47158	10-Nov-2023 04:56:54
43	XYZ1234	BLANK	UA	0.0031 -		47160	10-Nov-2023 04:57:30
44	XYZ1234	BLANK	UA	0.0030 -		47161	10-Nov-2023 04:57:48
45	XYZ1234	BLANK	UA	0.0031 -		47162	10-Nov-2023 04:58:06
46	2206133A	MULTICAL	UA	0.1120 -		47163	10-Nov-2023 04:58:24
47	2206133A	MULTICAL	UA	0.1119 -		47164	10-Nov-2023 04:58:42
48	2206133A	MULTICAL	UA	0.1124 -		47165	10-Nov-2023 04:59:01
49	XYZ1234	BLANK	ALPU	0.0003 -		47167	10-Nov-2023 07:23:45
50	XYZ1234	BLANK	ALPU	0.0004 -		47168	10-Nov-2023 07:24:03
51	XYZ1234	BLANK	ALPU	0.0005 -		47169	10-Nov-2023 07:24:21

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Nov-2023 To 30-Nov-2023

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
52	2206133A	MULTICAL	ALPU	0.1019 -		47170	10-Nov-2023 07:24:39
53	2206133A	MULTICAL	ALPU	0.1015 -		47171	10-Nov-2023 07:24:57
54	2206133A	MULTICAL	ALPU	0.1017 -		47172	10-Nov-2023 07:25:16
55	XYZ1234	BLANK	UA	0.0029 -		47174	10-Nov-2023 07:25:52
56	XYZ1234	BLANK	UA	0.0030 -		47175	10-Nov-2023 07:26:10
57	XYZ1234	BLANK	UA	0.0031 -		47176	10-Nov-2023 07:26:28
58	2206133A	MULTICAL	UA	0.1132 -		47177	10-Nov-2023 07:26:46
59	2206133A	MULTICAL	UA	0.1126 -		47178	10-Nov-2023 07:27:04
60	2206133A	MULTICAL	UA	0.1125 -		47179	10-Nov-2023 07:27:22
61	XYZ1234	BLANK	ALPU	0.0004 -		47181	10-Nov-2023 08:14:48
62	XYZ1234	BLANK	ALPU	0.0003 -		47182	10-Nov-2023 08:15:06
63	XYZ1234	BLANK	ALPU	0.0004 -		47183	10-Nov-2023 08:15:24
64	2206133A	MULTICAL	ALPU	0.1034 -		47184	10-Nov-2023 08:15:42
65	2206133A	MULTICAL	ALPU	0.1022 -		47185	10-Nov-2023 08:16:00
66	2206133A	MULTICAL	ALPU	0.1032 -		47186	10-Nov-2023 08:16:18
67	XYZ1234	BLANK	UA	0.0028 -		47188	10-Nov-2023 08:16:54
68	XYZ1234	BLANK	UA	0.0033 -		47189	10-Nov-2023 08:17:12

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Nov-2023 To 30-Nov-2023

Sr #	Lot #	Consumable	Test	Result	Unit	Flag	Curve #	Result Date
69	XYZ1234	BLANK	UA	0.0032	-		47190	10-Nov-2023 08:17:30
70	2206133A	MULTICAL	UA	0.1135	-		47191	10-Nov-2023 08:17:48
71	2206133A	MULTICAL	UA	0.1128	-		47192	10-Nov-2023 08:18:06
72	2206133A	MULTICAL	UA	0.1129	-		47193	10-Nov-2023 08:18:24
73	XYZ1234	BLANK	ALPU	NA	-	R2**	47327	11-Nov-2023 10:46:08
74	XYZ1234	BLANK	ALPU	NA	-	S**,R1**	47328	11-Nov-2023 10:46:26
75	XYZ1234	BLANK	ALPU	NA	-	S**,R1**	47329	11-Nov-2023 10:46:44
76	2206133A	MULTICAL	ALPU	NA	-	R1**	47330	11-Nov-2023 10:47:02
77	2206133A	MULTICAL	ALPU	NA	-	R2**	47331	11-Nov-2023 10:47:20
78	2206133A	MULTICAL	ALPU	NA	-	R2**	47332	11-Nov-2023 10:47:38
79	XYZ1234	BLANK	UA	NA	-	R2**	47334	11-Nov-2023 10:48:14
80	XYZ1234	BLANK	UA	NA	-	R2**	47335	11-Nov-2023 10:48:32
81	XYZ1234	BLANK	UA	0.0023	-		47336	11-Nov-2023 10:48:50
82	2206133A	MULTICAL	UA	0.0784	-		47337	11-Nov-2023 10:49:08
83	2206133A	MULTICAL	UA	0.0783	-		47338	11-Nov-2023 10:49:27
84	2206133A	MULTICAL	UA	0.0791	-		47339	11-Nov-2023 10:49:44
85	XYZ1234	BLANK	CHOL	0.0707	-		47511	13-Nov-2023 07:50:06

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Nov-2023 To 30-Nov-2023

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
86	XYZ1234	BLANK	CHOL	0.0709 -		47512	13-Nov-2023 07:50:24
87	XYZ1234	BLANK	CHOL	0.0684 -		47513	13-Nov-2023 07:50:42
88	2206133A	MULTICAL	CHOL	0.3285 -		47514	13-Nov-2023 07:51:00
89	2206133A	MULTICAL	CHOL	0.3283 -		47515	13-Nov-2023 07:51:18
90	2206133A	MULTICAL	CHOL	0.3300 -		47516	13-Nov-2023 07:51:36
91	XYZ1234	BLANK	UA	0.0029 -		47518	13-Nov-2023 07:52:12
92	XYZ1234	BLANK	UA	0.0034 -		47519	13-Nov-2023 07:52:30
93	XYZ1234	BLANK	UA	0.0034 -		47520	13-Nov-2023 07:52:48
94	2206133A	MULTICAL	UA	0.0793 -		47521	13-Nov-2023 07:53:06
95	2206133A	MULTICAL	UA	0.0802 -		47522	13-Nov-2023 07:53:24
96	2206133A	MULTICAL	UA	0.0790 -		47523	13-Nov-2023 07:53:42
97	XYZ1234	BLANK	UA	0.0042 -		47868	15-Nov-2023 18:19:19
98	XYZ1234	BLANK	UA	0.0041 -		47869	15-Nov-2023 18:19:37
99	XYZ1234	BLANK	UA	0.0016 -		47870	15-Nov-2023 18:19:55
100	2206133A	MULTICAL	UA	0.0730 -		47871	15-Nov-2023 18:20:13
101	2206133A	MULTICAL	UA	0.0717 -		47872	15-Nov-2023 18:20:31
102	2206133A	MULTICAL	UA	0.0719 -		47873	15-Nov-2023 18:20:50

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type : Calibration 01-Nov-2023 To 30-Nov-2023

Sr #	Lot #	Consumable	Test	Result	Unit	Flag	Curve #	Result Date
103	XYZ1234	BLANK	ALPU	0.0005	-		48328	20-Nov-2023 05:14:10
104	XYZ1234	BLANK	ALPU	0.0005	-		48329	20-Nov-2023 05:14:28
105	XYZ1234	BLANK	ALPU	0.0004	-		48330	20-Nov-2023 05:14:46
106	2206133A	MULTICAL	ALPU	0.0623	-		48331	20-Nov-2023 05:15:04
107	2206133A	MULTICAL	ALPU	0.0627	-		48332	20-Nov-2023 05:15:22
108	2206133A	MULTICAL	ALPU	0.0628	-		48333	20-Nov-2023 05:15:40
109	XYZ1234	BLANK	UA	0.0027	-	LV!	48335	20-Nov-2023 05:16:16
110	XYZ1234	BLANK	UA	0.0029	-		48336	20-Nov-2023 05:16:34
111	XYZ1234	BLANK	UA	0.0020	-		48337	20-Nov-2023 05:16:52
112	2206133A	MULTICAL	UA	0.0810	-		48338	20-Nov-2023 05:17:10
113	2206133A	MULTICAL	UA	0.0811	-		48339	20-Nov-2023 05:17:28
114	2206133A	MULTICAL	UA	0.0819	-		48340	20-Nov-2023 05:17:47
115	XYZ1234	BLANK	UA	0.0038	-		48697	24-Nov-2023 06:09:16
116	XYZ1234	BLANK	UA	0.0031	-		48698	24-Nov-2023 06:09:34
117	XYZ1234	BLANK	UA	0.0032	-		48699	24-Nov-2023 06:09:53
118	2206133A	MULTICAL	UA	0.0833	-		48700	24-Nov-2023 06:10:10
119	2206133A	MULTICAL	UA	0.0828	-		48701	24-Nov-2023 06:10:28

SP WELLFORT HOSPITAL

Ashrmam Lane, Sasthamangalam, Thiruvananthapuram

Result Reprint

Report Type

: Calibration

01-Nov-2023 To 30-Nov-2023

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date
120	2206133A	MULTICAL	UA	0.0823 -		48702	24-Nov-2023 06:10:46
121	XYZ1234	BLANK	UA	0.0021 -		49040	28-Nov-2023 12:35:13
122	XYZ1234	BLANK	UA	0.0026 -		49041	28-Nov-2023 12:35:31
123	XYZ1234	BLANK	UA	0.0029 -		49042	28-Nov-2023 12:35:49
124	2206133A	MULTICAL	UA	0.0823 -		49043	28-Nov-2023 12:36:07
125	2206133A	MULTICAL	UA	0.0836 -		49044	28-Nov-2023 12:36:25
126	2206133A	MULTICAL	UA	0.0823 -		49045	28-Nov-2023 12:36:43