



COMMITTED TO THE  
CUSTOMER SINCE - 1996

**VAIDYANATHESHWARA INSTRUMENTS**



CC-2473



**CERTIFICATE OF CALIBRATION**

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.  
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Page: 1 of 2

**1 Name and Address of the Customer**

M/s.: Acharya Tulsi Diagnostics Centre & Dental Care,  
# 228/12, MSP Plaza, Magadi Road,  
(Near Magadi Road Metro Station), Bangalore - 560 023.

**2 Customer Reference**

2.1 ULR No. : CC247323500124316F  
2.2 SRF No. : 3912  
2.3 Certificate No. : VI/23-24/3912-01  
2.4 Format No. : VI-FRM-TH-006  
2.5 Date of Issue : 20-01-2024

**3 Details Of Device Under Calibration(DUC).**

3.1 Nomenclature : **Temperature Indicator With Sensor of Refrigerator**  
3.2 Make / Model : Godrej / RTEON245B  
3.3 Id. No. : ATDC/REF-01  
3.4 DUC Condition : Satisfactory  
3.5 Calibration Procedure No. : SOP-38-06  
3.6 No.of Pages : 2  
3.7 Calibration Date : 17-01-2024  
3.8 Calibration Due : 16-01-2025  
3.9 Calibration done at : Onsite  
3.10 Instrument Location : Lab  
3.11 Discipline : Thermal

**4 Environmental Condition**

Temperature 23.2 - 24.4 °C Humidity 53 - 55 %RH

**5 Standards Used for Temperature Calibration**

Sl. No.	Nomenclature	Make / Model	Id. No.	Cal Agency	Certificate. No.	Validity
1	Paperless Recorder	Brain Child / VR-18	VI/OS/PR-12	VI, Bangalore	VI/22-23/INT-ET-428	28-03-2024
2	RTD Sensor	Frontier / ----	VI-RTD-02	VI Bangalore	VI/23-24/INT-TH-100-02	13-02-2024

**6 Note:**

- 6.1.The Calibration Certificate relates only to the above DUC
- 6.2.Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI.
- 6.3.Corrections/erasing, invalidate the Calibration Certificate.
- 6.4.Calibration of the DUC are traceable to National standards/International Standards
- 6.5.Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- 6.6.Results Reported are valid at the time of and under the stated conditions of measurements.
- 6.7.The usage of NABL Symbol is as per NABL guidelines given in NABL-133

Calibrated By

*Rakesh*  
Rakesh

(Calibration Engineer)

Checked By

*Umesh D*  
Umesh D

(Onsite In-Charge)





COMMITTED TO THE  
CUSTOMER SINCE - 1996



## CERTIFICATE OF CALIBRATION



CC-2473

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.  
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No. VI/23-24/3912-01

Page: 2 of 2

Results:

Freezer

Range / Resolution: -20 °C / 1 °C


Sl. No.	STD Reading (°C)	DUC Reading (°C)	Error Observed (°C)	Measurement Uncertainty ± (°C)	k Factor
1	-20.1	-20	0.1	0.732	2.0

Refrigerator

Range / Resolution: 2 to 8 °C / 1 °C

Sl. No.	STD Reading (°C)	DUC Reading (°C)	Error Observed (°C)	Measurement Uncertainty ± (°C)	k Factor
1	2.0	2	0.0	0.732	2.0
2	4.1	4	-0.1	0.732	2.0
3	6.1	6	-0.1	0.732	2.0
4	7.1	7	-0.1	0.732	2.0
5	8.2	8	-0.2	0.732	2.0

### Pictorial Representation of Sensors Placement



Note :

1. S Indicates Placement of Standard Sensor

Conclusion / Remarks:

- 1 Measurement Uncertainty reported is at 95.45% confidence level with k = 2.

Calibrated By

*Rakesh*  
Rakesh

(Calibration Engineer)

Checked By

*Urvesh D*  
Urvesh D

(Onsite In-Charge)

