



CALIBRATION CERTIFICATE

ULR NO:-CC325723000001127F

M/S:- Delhi Diagnostics Main Lab 28-29 Chowk, old Faridabad	Service request No & date	KL/157-19/10/23
	Date of receiving	19/10/2023
	Date of calibration	21/10/2023
	Suggested Due Date	20/10/2024
	Date of issue	21/10/2023

Instrument Details			
Instrument Name	Micro Pipette	Sr.No.	257225
Make/Model	-	Location	-----
Range / Size	10 to 100 µl	I.D. No.	-----
Least Count	5 µl	Visual Inspection	OK
Calibration Done	At Lab	If Any Repair/Adjustment	NO

Detail of reference standards & Major equipments used					
Standards	Make	Serial No.	Certificate No	Next Due Date	Traceability
Dig. Weighing Balance	ACZET	CHA2204245	CC322623000000458F	16/04/2024	Tycon

Environmental Condition	Temperature	23±3 °C	Calibration Reference	IS:4162 (1985)
	Relative Humidity	50±10%	Calibration Procedure	WI:45

Calibration Results

Serial No.	Nominal Capacity in (µl)	Measured Capacity in (µl)	Uncertainty At 95% C.L. (coverage factor k=2)
01.	20.0	19.97585	±0.9 µl
02.	50.0	49.86957	±0.9 µl
03.	100.0	99.87846	±0.9 µl

REMARKS:-

- NOTE a) This result of calibration refers to only to the particular item submitted for calibration.
 b) The above results are valid at the time of and under the stated conditions measurement.
 c) This report should not be reproduced except in full without the written permission of CEO.
 d) The traceability used in calibration is traceable to national / international standards.

Calibrated By
 (Calibration/Wing./TM)
 Page 1 of 1

Approved By
 (QM/TM)
 (C.S.MAHAPATRA)

