



## CALIBRATION CERTIFICATE

ULR NO:-CC325723000001128F

M/S:- Delhi Diagnostics Main Lab 28-29 Chowk, old Faridabad	Service request No & date	KL/157-19/10/23
	Date of receiving	19/10/2023
	Date of calibration	21/10/2023
	Suggested Due Date	20/10/2024
	Date of issue	21/10/2023

### Instrument Details

Instrument Name	Micro Pipette	Sr.No.	052133322
Make/Model	-	Location	-----
Range / Size	100 to 1000 µl	I.D. No.	-----
Least Count	10 µl	Visual Inspection	OK
Calibration Done	At Lab	If Any Repair/Adjustment	NO

### Detail of reference standards & Major equipments used

Standards	Make	Serial No.	Certificate No	Next Due Date	Traceability
Dig. Weighing Balance	ACZET	CHA2204245	CC322623000000458F	16/04/2024	Tycon

Environmental Condition	Temperature	23±3 °C	Calibration Reference	IS:4162 (1985)
	Relative Humidity	50±10%	Calibration Procedure	WI:45

### Calibration Results

Serial No.	Nominal Capacity in (µl)	Measured Capacity in (µl)	Uncertainty At 95% C.L. (coverage factor k=2)
01.	100	99.88	±0.9 µl
02.	500	499.79	
03.	1000	999.65	

### REMARKS:-

- NOTE a) This result of calibration refers to only to the particular item submitted for calibration.  
 b) The above results are valid at the time of and under the stated conditions measurement.  
 c) This report should not be reproduced except in full without the written permission of CEO.  
 d) The traceability used in calibration is traceable to national / international standards.

Calibrated By  
(Calibration Engg./TM)  
Page 1 of 1



-----End of Report-----

Approved By  
(QM/TM)  
(C.S.MAHAPATRA)