

HORIBA

Explore the future

HORIBA India Private Ltd.

246, Okhla Industrial

Estate, Phase - III 110

020 New Delhi, India

Tel: +91 (11) 4646 5000 / 4669 5001

Fax: +91 (11) 4669 5010 / 4646 5020

HIN/MED/2024-2025/0380

10th February 2024.

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer ABX Micros ES 60 bearing serial number: **703ESOH11085** installed **SIDDHAM PATHOLOGY LABORATORY DIAGNOSTICS SPL DIAGNOSTICS** Biopal was Calibrated on **10th February 2024.**

Calibrator : ABX MINOCAL

Lot No. : CX491

Expiry Date : 05th March 2024.

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 09th February 2025.



Shrish Dixit
(Head- Products & Marketing)
For **Horiba India Pvt. Ltd.**



Explore the Future

Automotive Test Systems | Process & Environment | Medical | Semiconductor | Scientific

HORIBA
Explore the Future