



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - OCTOBER 2023



PC-1024

Lab Name **ANUSHKA DIAGNOSTICS** Lab No **19709**
 Constituent Group **Chemistry II** Date of Result Entered : **18/10/2023**
 PT item **Lyophilized human serum based** Date of Report Published : **04/11/2023**

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of		Participants		Your Value	SDI	U
				Participants	DV	CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	232	363.43	5.06	18.40	359 mg/dL	-0.24	2.42
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	276	43.46	6.10	2.65	43.2 mg/dL	-0.10	0.32
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	106	7.69	5.78	0.44	6.54 mg/dL	-2.59	0.09
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	243	1.36	12.45	0.17	1.35 mg/dL	-0.06	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	292	4.90	6.57	0.32	4.82 g/DI	-0.25	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	283	3.10	6.62	0.20	2.67 g/dL	-2.10	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	285	10.23	6.26	0.64	9.44 mg/dL	-1.23	0.08
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	290	101.59	6.61	6.71	95 mg/dL	-0.98	0.79
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	270	80.28	9.41	7.55	82 mg/dL	0.23	0.92
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
ANUSHKA DIAGNOSTICS
ANUSHKA LIFE CARE PVT LTD, 115/102, STATION ROAD,
BOLPUR
WEST BENGAL731204

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****