



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - DECEMBER 2023



PC-1024

Lab Name ANUSHKA DIAGNOSTICS Lab No 19709  
Constituent Group Chemistry II Date of Result Entered : 12/12/2023  
PT Item Lyophilized human serum based Date of Report Published : 02/01/2024

Sl No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	240	203.83	5.26	10.72	213 mg/dL	0.86	1.38
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	290	111.33	7.18	8.00	108.15 mg/dL	-0.40	0.94
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	109	8.03	4.97	0.40	7.49 mg/dL	-1.35	0.08
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	243	2.00	11.15	0.22	1.97 mg/dL	-0.13	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	288	5.40	7.00	0.38	5.31 g/dL	-0.24	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	283	3.38	6.24	0.21	3.22 g/dL	-0.76	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	302	6.84	7.95	0.54	6.44 mg/dL	-0.74	0.06
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	307	119.92	6.73	8.07	108 mg/dL	-1.48	0.92
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	305	124.11	8.62	10.70	122 mg/dL	-0.20	1.23

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

LAB ADDRESS :  
ANUSHKA DIAGNOSTICS  
ANUSHKA LIFE CARE PVT LTD, 115/102, STATION ROAD,  
BOLPUR  
WEST BENGAL 731204

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*