

Calibration Certificate

This is to certify that the below said instrument was calibrated on 01/04/2023 at WellCare Clinical Lab , Zirakpur and found working satisfactory.

Calibration done on 01/04/2023

Next calibration due on

31/03/2024

Serial No : 3020ET03772

Instrument : Automated Hematology

Analyzer

Model : Cellenium 21

Calibration Done by : PARSHANT

Calibration Parameters:

- 1) Input Power -229v
- 2) Earthing -0.2v
- 3) Display Checked –OK
- 4) Keypad checked –OK
- 5) HGB Blank voltage checked – OK 4.19v.

The Automated Hematology Analyzer Cellenium 21 has been calibrated with calibrator Lot No : PLUS1117, EXP date: 12/10/2024 & found working satisfactory.

For Trivitron Healthcare Pvt Ltd,

Parshant

Assistant Manager Service - Chandigarh

Trivitron Healthcare Pvt. Ltd.

#15,IVStreet,Abhiramapuram,Chennai-600018,India.

Ph: +91 - 44 - 2498 5050. Fax: +91 - 44 - 2498 5757 / 2467 / 2782

E-mail: trivitron@trivitron.com

www.trivitron.com

BANGALORE(+9180653489988)|DELHI(+911141354650)|HYDERABAD(+914065618910)|KOLKATA(+913324650298/0469)|MUMBAI(+912266612526/27/28, 6610 1589/90)

BranchOffices:Ahmedabad,Chandigarh,Cochin,Coimbatore,Jaipur,Lucknow,Nagpur,Pondicherry,Pune,SriLanka&Dubai

