

SelfTest Dec 2023
 Thursday, December 21, 2023, 13:43

Lamp - OK

Pipetting Accuracy: Small Sample Medium Sample Large Sample
 photometer Repeatability: OK OK OK

PASSED

DETAILS:
 Version 1.2.0.329//1086
 (00.15.56/V 00.15.56/X 00.15.56/Z 00.15.56/P 00.15.56)

Temperature at start: 25.7 °C Temperature at end: 25.7 °C

Filter:	340	405	505	545	580	630	000	000
rVolts:	5.706	5.973	6.203	6.409	6.457	6.654	0.002	0.002
PTVolts:	3.155	2.876	3.792	4.064	3.950	4.376	0.002	0.002

Type	Well	Readings (405/630)			Mean	SD	%CV
Blank	A1	0.0115	0.0115	0.0116	0.0115	0.0001	
	A2	0.0133	0.0146	0.0135	0.0138	0.0007	
Small	A3	0.3913	0.3915	0.3921	0.3916	0.0004	0.1063
	A4	0.4005	0.4013	0.4007	0.4008	0.0004	0.1039
	A5	0.3940	0.3949	0.3948	0.3946	0.0005	0.1250
	B1	0.3915	0.3915	0.3920	0.3917	0.0003	0.0737
	B2	0.3935	0.3935	0.3941	0.3937	0.0003	0.0880
	B3	0.3946	0.3944	0.3950	0.3947	0.0003	0.0774
Medium	B4	1.7030	1.7021	1.7050	1.7034	0.0015	0.0871
	B5	1.7019	1.7035	1.7057	1.7037	0.0019	0.1120
	C1	1.7021	1.7051	1.7044	1.7039	0.0016	0.0921
	C2	1.7059	1.7088	1.7092	1.7080	0.0018	0.1054
	C3	1.7087	1.7097	1.7087	1.7090	0.0006	0.0338
Large	C4	1.7041	1.7043	1.7048	1.7044	0.0004	0.0212
	C5	2.5166	2.5163	2.5196	2.5175	0.0018	0.0725
	D1	2.5223	2.5229	2.5287	2.5246	0.0035	0.1400
	D2	2.5899	2.5946	2.5990	2.5945	0.0046	0.1754
	D3	2.5977	2.5971	2.6022	2.5990	0.0028	0.1073
	D4	2.6006	2.6018	2.5990	2.6005	0.0014	0.0540
	D5	2.5940	2.5979	2.5980	2.5966	0.0023	0.0878

Type	Target	Mean	SD	%CV	%Diff
Large	2.6720	2.5721	0.0373	1.4517	-3.7379
Medium	1.7300	1.7054	0.0026	0.1522	-1.4226
Small	0.3560	0.3945	0.0032	0.8083	10.8177