



Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092

Ph.: 011-42657237 Mob.: 9810804449, 8368671530

E- mail: primac21@gmail.com



Format No.
F01(7.8/01)

CALIBRATION CERTIFICATE

ULR No. : CC233924000000304F	Calibration Done on : 24.01.2024	S.R.F. No. : 0002/24
Issue Date : 27.01.2024	Calibration Valid Upto : 23.01.2025	S.R.F. Date : 24.01.2024
UUC Receipt Date : N/A	Certificate No. : PCS/0304	Page No. : 1 of 1

Calibrated for M/S.

Lupin Diagnostics Ltd.
Ground Floor SRP Arcade, E-5/48, Opp Ajwani Eye Care Center
Arera Colony, Vittal Market, Bhopal (M.P.) 462016

Identification on UUC (Unit Under Calibration)

Instrument Name	Centrifuge	Size/Range	(0 - 5250) RPM
Least Count	10 RPM	Condition of Item	Good
Manufactured By	Remi	Calibration Performed at	Site
Manufacturer M/Sr. No.	R-8C / ZIIN-34967	Location	Lab
Party ID Mark	LDL-BPL/CEN-01	Ref. Standard / Method	IS: 12508
		Acceptance Criteria	N/A

Calibration Procedure : As per Calibration Procedure Document No. PCS/CP/01, Section No.CP-S-04

Environmental conditions (s) : Temperature : (25 ± 10) °C. Relative Humidity : (50 ± 15) % RH

Standard & Major Equipment(s)

Standard(s) & Equipments	ID No. / Serial No.	Calibrating Agency	Traceability ULR No. / Cert. No.	Valid Up to
Dig. Tachometer	PCS/DTMC/01 / 109-0595Y	Transcal	TSC/22-23/18608-1	14.03.2024

Calibration Results:

UUC Value (RPM)	Avg. Std. Value (RPM)	Error (RPM)	Uncertainty of measurement (at Approx 95% Confidence Level With Coverage Factor k=2) = ±
100	100.6	-0.6	1.60%
500	500.9	-0.9	1.60%
1000	1001.2	-1.2	1.60%
3000	3002.8	-2.8	1.60%
4500	4503.4	-3.4	1.60%

Remark :-

CALIBRATED / CHECKED BY

(CALIBRATION ENGINEER)



APPROVED BY

AFTAB AHAMAD
DY. (QM/TM)

***** END OF CERTIFICATE *****



Note : The Certificate has issued subject to conditions stated back side overleaf



Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092
Ph.: 011-42657237 Mob.: 9810804449, 8368671530
E- mail: primac21@gmail.com



CC-2339

Format No.
F01(7.8/01)

CALIBRATION CERTIFICATE

ULR No. : CC23392400000307F	Calibration Done on : 24.01.2024	S.R.F. No. : 0001/07
Issue Date : 27.01.2024	Calibration Valid Upto : 23.01.2025	S.R.F. Date : 07.07.2023
UUC Receipt Date : N/A	Certificate No. : PCS/0307	Page No. : 01 of 01
Calibrated for M/S. Lupin Diagnostics Ltd. Ground Floor SRP Arcade, E-5/48, Opp Ajwani Eye Care Center Arera Colony, Vittal Market, Bhopal (M.P.) 462016		

Identification on UUC (Unit Under Calibration)

Instrument Name	Dig. Timer	Size/Range	(0 to 99) Min
Least Count	1 Min	Condition of Item	Good
Manufactured By	Remi	Calibration Performed at	Site
Manufacturer M/Sr.No.	R-8C / ZIIN-34967	Location	Lab
Party ID Mark	LDL-BPL/CEN-01	Ref. Standard / Method	NIST 960-12
		Acceptance Criteria	N/A

Calibration Procedure : As per Calibration Procedure Document No. PCS/CPET/03, Section No.CPET-S-10(S)

Environmental Conditions : Temperature : (25 ± 10)°C. Relative Humidity : (50 ± 20) % RH

Standard & Major Equipment(s) Used :

Standard(s) & Equipments	ID No. / Serial No.	Calibrating Agency	Traceability ULR No. / Cert. No.	Valid Up to
Digital Time Calibrator	PCS/DTM/01 / 34452101	Transcal	TSC/22-23/18608-6	15.03.2024

Calibration Results :

UUC Value (in Sec.)	Avg. Std. Value (in Sec.)	Error (in Sec.)	Related Uncertainty on C.F.(k=2) at approx 95% Confidence Level ± (Sec.)
60	60.048	-0.048	34.6
600	600.13	-0.13	34.6
1200	1200.21	-0.21	34.7

Remark : --

CALIBRATED / CHECKED BY

(CALIBRATION ENGINEER)



APPROVED BY

AFTAB AHAMAD
DY.(QM/TM)

***** END OF CERTIFICATE *****



Note : The Certificate has issued subject to conditions stated back side overleaf



Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092
Ph.: 011-42657237 Mob.: 9810804449, 8368671530
E- mail: primac21@gmail.com



CC-2339

Format No.
F01(7.8/01)

CALIBRATION CERTIFICATE

ULR No. : CC233924000000305F	Calibration Done on : 24.01.2024	S.R.F. No. : 0002/24
Issue Date : 27.01.2024	Calibration Valid Upto : 23.01.2025	S.R.F. Date : 24.01.2024
UUC Receipt Date : N/A	Certificate No. : PCS/0305	Page No. : 1 of 1
Calibrated for M/S. Lupin Diagnostics Ltd. Ground Floor SRP Arcade, E-5/48, Opp Ajwani Eye Care Center Arera Colony, Vittal Market, Bhopal (M.P.) 462016		

Identification on UUC (Unit Under Calibration)

Instrument Name	Centrifuge	Size/Range	(0 - 5250) RPM
Least Count	10 RPM	Condition of Item	Good
Manufactured By	Remi	Calibration Performed at	Site
Manufacturer M/Sr. No.	R-8C / ZIIN-34968	Location	Lab
Party ID Mark	LDL-BPL/CEN-02	Ref. Standard / Method	IS: 12508
		Acceptance Criteria	N/A

Calibration Procedure : As per Calibration Procedure Document No. PCS/CP/01, Section No.CP-S-04

Environmental conditions (s) : Temperature : (25 ± 10) °C. Relative Humidity : (50 ± 15) % RH

Standard & Major Equipment(s)

Standard(s) & Equipments	ID No. / Serial No.	Calibrating Agency	Traceability ULR No. / Cert. No.	Valid Up to
Dig. Tachometer	PCS/DTMC/01 / 109-0595Y	Transcal	TSC/22-23/18608-1	14.03.2024

Calibration Results:

UUC Value (RPM)	Avg. Std. Value (RPM)	Error (RPM)	Uncertainty of measurement (at Approx 95% Confidence Level With Coverage Factor k=2) = ±
100	99.8	0.2	1.60%
500	499.6	0.4	1.60%
1000	1001.6	-1.6	1.60%
3000	3002.4	-2.4	1.60%
4500	4504.1	-4.1	1.60%

Remark :--

CALIBRATED / CHECKED BY

(CALIBRATION ENGINEER)



APPROVED BY

AFTAB AHAMAD
DY. (QM/TM)

***** END OF CERTIFICATE *****

Note : The Certificate has issued subject to conditions stated back side overleaf





Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092

Ph.: 011-42657237 Mob.: 9810804449, 8368671530

E- mail: primac21@gmail.com



Format No.
F01(7.8/01)

CALIBRATION CERTIFICATE

ULR No. : CC23392400000308F	Calibration Done on : 24.01.2024	S.R.F. No. : 0001/07
Issue Date : 27.01.2024	Calibration Valid Upto : 23.01.2025	S.R.F. Date : 07.07.2023
UUC Receipt Date : N/A	Certificate No. : PCS/0308	Page No. : 01 of 01
Calibrated for M/S. Lupin Diagnostics Ltd. Ground Floor SRP Arcade, E-5/48, Opp Ajwani Eye Care Center Arera Colony, Vittal Market, Bhopal (M.P.) 462016		

Identification on UUC (Unit Under Calibration)

Instrument Name	Dig. Timer	Size/Range	(0 to 99) Min
Least Count	1 Min	Condition of Item	Good
Manufactured By	Remi	Calibration Performed at	Site
Manufacturer M/Sr.No.	R-8C / ZIIN-34968	Location	Lab
Party ID Mark	LDL-BPL/CEN-02	Ref. Standard / Method	NIST 960-12
		Acceptance Criteria	N/A

Calibration Procedure : As per Calibration Procedure Document No. PCS/CPET/03, Section No.CPET-S-10(S)

Environmental Conditions : Temperature : (25 ± 10)°C. Relative Humidity : (50 ± 20) % RH

Standard & Major Equipment(s) Used :

Standard(s) & Equipments	ID No. / Serial No.	Calibrating Agency	Traceability ULR No. / Cert. No.	Valid Up to
Digital Time Calibrator	PCS/DTM/01 / 34452101	Transcal	TSC/22-23/18608-6	15.03.2024

Calibration Results :

UUC Value (in Sec.)	Avg. Std. Value (in Sec.)	Error (in Sec.)	Related Uncertainty on C.F.(k=2) at approx 95% Confidence Level ± (Sec.)
60	60.067	-0.067	34.6
600	600.19	-0.19	34.6
1200	1200.42	-0.42	34.7

Remark : --

CALIBRATED / CHECKED BY

(CALIBRATION ENGINEER)



APPROVED BY

AFTAB AHAMAD
DY.(QM/TM)

***** END OF CERTIFICATE *****



Note : The Certificate has issued subject to conditions stated back side overleaf



Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092
Ph.: 011-42657237 Mob.: 9810804449, 8368671530
E- mail: primac21@gmail.com



Format No.
F01(7.8/01)

CALIBRATION CERTIFICATE

ULR No. : CC233924000000306F	Calibration Done on : 24.01.2024	S.R.F. No. : 0002/24
Issue Date : 27.01.2024	Calibration Valid Upto : 23.01.2025	S.R.F. Date : 24.01.2024
UUC Receipt Date : N/A	Certificate No. : PCS/0306	Page No. : 1 of 1
Calibrated for M/S. Lupin Diagnostics Ltd. Ground Floor SRP Arcade, E-5/48, Opp Ajwani Eye Care Center Arera Colony, Vittal Market, Bhopal (M.P.) 462016		

Identification on UUC (Unit Under Calibration)

Instrument Name	Centrifuge	Size/Range	(0 - 5250) RPM
Least Count	10 RPM	Condition of Item	Good
Manufactured By	Remi	Calibration Performed at	Site
Manufacturer M/Sr. No.	R-8C / ZIIN-34965	Location	Lab
Party ID Mark	LDL-BPL/CEN-03	Ref. Standard / Method	IS: 12508
		Acceptance Criteria	N/A

Calibration Procedure : As per Calibration Procedure Document No. PCS/CP/01, Section No.CP-S-04

Environmental conditions (s) : Temperature : (25 ± 10) °C. Relative Humidity : (50 ± 15) % RH

Standard & Major Equipment(s)

Standard(s) & Equipments	ID No. / Serial No.	Calibrating Agency	Traceability ULR No. / Cert. No.	Valid Up to
Dig. Tachometer	PCS/DTMC/01 / 109-0595Y	Transcal	TSC/22-23/18608-1	14.03.2024

Calibration Results:

UUC Value (RPM)	Avg. Std. Value (RPM)	Error (RPM)	Uncertainty of measurement (at Approx 95% Confidence Level With Coverage Factor k=2) = ±
100	100.2	-0.2	1.60%
500	500.8	-0.8	1.60%
1000	1001.7	-1.7	1.60%
3000	3002.8	-2.8	1.60%
4500	4504.3	-4.3	1.60%

Remark :-

CALIBRATED / CHECKED BY

(CALIBRATION ENGINEER)



APPROVED BY

AFTAB AHAMAD
DY. (QM/TM)

***** END OF CERTIFICATE *****



Note : The Certificate has issued subject to conditions stated back side overleaf



Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092

Ph.: 011-42657237 Mob.: 9810804449, 8368671530

E- mail: primac21@gmail.com



Format No.
F01(7.8/01)

CALIBRATION CERTIFICATE

ULR No. : CC233924000000309F	Calibration Done on : 24.01.2024	S.R.F. No. : 0001/07
Issue Date : 27.01.2024	Calibration Valid Upto : 23.01.2025	S.R.F. Date : 07.07.2023
UUC Receipt Date : N/A	Certificate No. : PCS/0309	Page No. : 01 of 01
Calibrated for M/S. Lupin Diagnostics Ltd. Ground Floor SRP Arcade, E-5/48, Opp Ajwani Eye Care Center Arera Colony, Vittal Market, Bhopal (M.P.) 462016		

Identification on UUC (Unit Under Calibration)

Instrument Name	Dig. Timer	Size/Range	(0 to 99) Min
Least Count	1 Min	Condition of Item	Good
Manufactured By	Remi	Calibration Performed at	Site
Manufacturer M/Sr.No.	R-8C / ZIIN-34965	Location	Lab
Party ID Mark	LDL-BPL/CEN-03	Ref. Standard / Method	NIST 960-12
		Acceptance Criteria	N/A

Calibration Procedure : As per Calibration Procedure Document No. PCS/CPET/03, Section No. CPET-S-10(S)

Environmental Conditions : Temperature : (25 ± 10)°C. Relative Humidity : (50 ± 20) % RH

Standard & Major Equipment(s) Used :

Standard(s) & Equipments	ID No. / Serial No.	Calibrating Agency	Traceability ULR No. / Cert. No.	Valid Up to
Digital Time Calibrator	PCS/DTM/01 / 34452101	Transcal	TSC/22-23/18608-6	15.03.2024

Calibration Results :

UUC Value (in Sec.)	Avg. Std. Value (in Sec.)	Error (in Sec.)	Related Uncertainty on C.F.(k=2) at approx 95% Confidence Level ± (Sec.)
60	60.081	-0.081	34.6
600	600.12	-0.12	34.6
1200	1200.24	-0.24	34.7

Remark : --

CALIBRATED / CHECKED BY

(CALIBRATION ENGINEER)



APPROVED BY

AFTAB AHAMAD
DY.(QM/TM)

***** END OF CERTIFICATE *****



Note : The Certificate has issued subject to conditions stated back side overleaf