



Prima Calibration Services

F-199, MAᅅGAL BAZAR LAXMI NAGAR DELHI-110092
Ph.: 011-42657237 Mob.: 9810804449, 8368671530
E- mail: primac21@gmail.com



CC-2339

Format No.
F01(7.8/01)

CALIBRATION CERTIFICATE

ULR No. : CC233924000000314F	Calibration Done on : 25.01.2024	S.R.F. No. : 0002/24
Issue Date : 27.01.2024	Calibration Valid Upto : 24.01.2025	S.R.F. Date : 24.01.2024
UUC Receipt Date : 24.01.2024	Certificate No. : PCS/0314	Page No. : 1 of 1

Calibrated for M/S.
Lupin Diagnostics Ltd.
Ground Floor SRP Arcade, E-5/48,Opp Ajwani Eye Care Center
Arera Colony, Vittal Market, Bhopal (M.P.) 462016

Identification on UUC (Unit Under Calibration)

Instrument Name	Micro Pipette (Piston Type)	Size/Range	(2 - 20) μ l
Least Count	0.1 μ l	Condition of Item	Good
Manufactured By	MICROLIT	Calibration Performed at	Lab
Manufacturer M/S.No.	--/22408440	Location	N/A
Party ID Mark	LDL-BPL/MP-01	Ref. Standard / Method	ISO-4787
		Acceptance Criteria	N/A

Calibration Procedure : As per Calibration Procedure Document No. PCS/CP/01, Section No.CP-50

Environmental conditions (s) : Temperature: (23 \pm 3) $^{\circ}$ C. Relative Humidity : (50 \pm 10)% RH

Standard & Major Equipment(s) Used :

Standard(s) & Equipments	ID No. / Serial No.	Calibrating Agency	Traceability ULR No. / Cert. No.	Valid Up to
Dig. Weighing Balance (Upto 200 g)	PCS/DWB/01 / 529818	BLUBOX TECHNO	CC308223000000042F	14.02.2024

Calibration Results at 27 $^{\circ}$ C :

Nominal Value UUC (μ l)	Avg. Observed Value (μ l)	Error (μ l)	Related Uncertainty on C.F.(k=2) at approx 95% Confidence Level \pm (μ l)
10.0	9.93	0.07	0.13
20.0	19.75	0.25	0.13

Remark :---

CALIBRATED / CHECKED BY

(CALIBRATION ENGINEER)



APPROVED BY

AFTAB AHAMAD
DY. (QM/TM)

***** END OF CERTIFICATE *****



Note : The Certificate has issued subject to conditions stated back side overleaf



Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092

Ph.: 011-42657237 Mob.: 9810804449, 8368671530

E- mail: primac21@gmail.com



CC-2339

Format No.
F01(7.8/01)

CALIBRATION CERTIFICATE

ULR No. : CC233924000000316F	Calibration Done on : 25.01.2024	S.R.F. No. : 0002/24
Issue Date : 27.01.2024	Calibration Valid Upto : 24.01.2025	S.R.F. Date : 24.01.2024
UUC Receipt Date : 24.01.2024	Certificate No. : PCS/0316	Page No. : 1 of 1

Calibrated for M/S.
Lupin Diagnostics Ltd.
Ground Floor SRP Arcade, E-5/48, Opp Ajwani Eye Care Center
Arera Colony, Vittal Market, Bhopal (M.P.) 462016

Identification on UUC (Unit Under Calibration)

Instrument Name	Micro Pipette (Piston Type)	Size/Range	(20-200) µl
Least Count	1 µl	Condition of Item	Good
Manufactured By	MICROLIT	Calibration Performed at	Lab
Manufacturer M/S.No.	--/TW11287	Location	N/A
Party ID Mark	LDL-BPL/MP-02	Ref. Standard / Method	ISO-4787
		Acceptance Criteria	N/A

Calibration Procedure : As per Calibration Procedure Document No. PCS/CP/01, Section No.CP-50

Environmental conditions (s) : Temperature: (23 ± 3) °C. Relative Humidity : (50 ± 10)% RH

Standard & Major Equipment(s) Used :

Standard(s) & Equipments	ID No. / Serial No.	Calibrating Agency	Traceability ULR No. / Cert. No.	Valid Up to
Dig. Weighing Balance (Upto 200 g)	PCS/DWB/01 / 529818	BLUBOX TECHNO	CC308223000000042F	14.02.2024

Calibration Results at 27 °C :

Nominal Value UUC (µl)	Avg. Observed Value (µl)	Error (µl)	Related Uncertainty on C.F.(k=2) at approx 95% Confidence Level ± (µl)
20	20.05	-0.05	0.13
100	99.98	0.02	0.56
200	200.19	-0.19	0.56

Remark :---

CALIBRATED / CHECKED BY

(CALIBRATION ENGINEER)



APPROVED BY

AFTAB AHAMAD
DY. (QM/TM)

***** END OF CERTIFICATE *****



Note : The Certificate has issued subject to conditions stated back side overleaf



Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092
Ph.: 011-42657237 Mob.: 9810804449, 8368671530
E- mail: primac21@gmail.com



Format No.
F01(7.8/01)

CALIBRATION CERTIFICATE

ULR No. : CC23392400000317F	Calibration Done on : 25.01.2024	S.R.F. No. : 0002/24
Issue Date : 27.01.2024	Calibration Valid Upto : 24.01.2025	S.R.F. Date : 24.01.2024
UUC Receipt Date : 24.01.2024	Certificate No. : PCS/0317	Page No. : 1 of 1

Calibrated for M/S.
Lupin Diagnostics Ltd.
Ground Floor SRP Arcade, E-5/48, Opp Ajwani Eye Care Center
Anera Colony, Vittal Market, Bhopal (M.P.) 462016

Identification on UUC (Unit Under Calibration)

Instrument Name	Micro Pipette (Piston Type)	Size/Range	(100-1000) µl
Least Count	5 µl	Condition of Item	Good
Manufactured By	MICROLIT	Calibration Performed at	Lab
Manufacturer M/S.No.	--/22400186	Location	N/A
Party ID Mark	LDL-BPL/MP-03	Ref. Standard / Method	ISO-8655-6
		Acceptance Criteria	N/A

Calibration Procedure : As per Calibration Procedure Document No. PCS/CP/01, Section No.CP-50

Environmental conditions (s) : Temperature: (23 ± 3) °C. Relative Humidity : (50 ± 10)% RH

Standard & Major Equipment(s) Used :

Standard(s) & Equipments	ID No. / Serial No.	Calibrating Agency	Traceability ULR No. / Cert. No.	Valid Up to
Dig. Weighing Balance (Upto 200 g)	PCS/DWB/01 / 529818	BLUBOX TECHNO	CC308223000000042F	14.02.2024

Calibration Results at 27 °C :

Nominal Value UUC (µl)	Avg. Observed Value (µl)	Error (µl)	Related Uncertainty on C.F.(k=2) at approx 95% Confidence Level ± (µl)
100	99.57	0.43	0.56
500	498.75	1.25	0.56
1000	997.63	2.37	0.56

Remark :---

CALIBRATED / CHECKED BY

(CALIBRATION ENGINEER)



APPROVED BY

AFTAB AHAMAD
DY. (QM/TM)

***** END OF CERTIFICATE *****



Note : The Certificate has issued subject to conditions stated back side overleaf