

SID:CREL2PRE CREATININE 1.50 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL2PRE CREATININE 1.52 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL2PRE CREATININE 1.48 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL2PRE CREATININE 1.52 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL2PRE CREATININE 1.49 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE 3.83 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE 3.79 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE 3.84 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE 3.85 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE 3.85 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE 3.89 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE 3.87 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE 3.91 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE	ID:	Date:06/22/2023R#: 4

J. M. W.

SID:CREL2PRE CREATININE 1.50 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL2PRE CREATININE 1.52 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL2PRE CREATININE 1.49 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL2PRE CREATININE 1.49 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL2PRE CREATININE 1.53 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL2PRE CREATININE 1.52 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL2PRE CREATININE 1.54 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL2PRE CREATININE 1.53 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL2PRE CREATININE 1.53 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL2PRE CREATININE 1.53 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL2PRE CREATININE 1.52 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL2PRE CREATININE 1.52 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL2PRE CREATININE 1.53 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL2PRE CREATININE 1.50 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL2PRE CREATININE	ID:	Date:06/22/2023R#: 4





Rx Daytona +

Calibration –Certificate –Rx Daytona +

Instrument	Rx Daytona + Clinical chemistry Analyzer
Serial No	7241-0663
Account Name	CluePath Diagnostic
Installation Date	19 Jun 2023
Calibration Date	19 Jun 2023
Next Calibration	18 Jun 2024

This is to certify that this analyzer has been inspected and calibration for following parameters

Test Parameter	Target value	Obtained value
Input Voltage	230-240 Volts	232 Volts
Incubator Temperature	37 ° C ± 0.3 Max	37 ° C ± 0.14
Reagent Tray Temperature	8-15 ° C	7.87 °C
Detector Performance Check All wavelengths	>8.3 Volts	~9.00 Volts
12V Lamp Supply	12 ± 0.3 volts	11.96 Volts
5V Supply	5 ± 0.3 volts	5.05 volts
24V Supply	24 ± 0.3 volts	24.03 volts
DTR Position	Between -20 & +20	Between -15 & +20
Water Blank	<5000	<1000

The results obtained are as per specifications & tolerance ranges. The above calibration was done with a Digital Multimeter (Model: MAS830L, Serial Number: 404994) and Digital Thermometer (Model: Lutron/TM-917, Serial Number G.51771) Calibration certificate is enclosed.

For Randox Laboratories India Pvt Ltd


Authorized Signatory
Samson Sundarra
Customer Support Engineer





Rx Daytona +

PERFORMANCE QUALIFICATION

FOR

**CluePath Diagnostic
19, New High School Road, V V Puram
Bangalore 560004**

RANDOX LABORATORIES LIMITED



PERFORMANCE QUALIFICATION

System Certification

Study data has been determined that the system described in this document either meets all criterial outlines in this Installation Qualification, or exceptional conditions have been identified and documentation included.
Exceptional conditions, if any have been addressed.
The system is ready for specified usage.

Protocol performed by: Randox Laboratories Representative
Name: Samson Sundarraj
Title: Customer Support Engineer
Company: Randox laboratories India Pvt Ltd.,

Customer authorization:
Name: Dr. MSN Prasad
Title: Sr. Pathologist & Managing Partner
CluePath Diagnostic
19, New High School Road, V V Puram
Bangalore 560004

J. Anurag

Engineer Name & Signature
Samson Sundarraj
Date: 19 Jun 2023



Customer Signature
Dr. MSN Prasad
Date: 19 Jun 2023

SID:CREL3PRE CREATININE 3.86 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE 3.89 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE 3.88 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE 3.85 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE 3.94 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE 3.91 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE 3.83 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE 3.97 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE 3.86 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE 3.88 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE 3.89 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE 3.89 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:CREL3PRE CREATININE 3.95 MG/DL 569 570	ID:	Date:06/22/2023R#: 4
SID:97000001 IRON	ID:RX L2 1584UN	S Lot:1584 UNDate:06/22/2023R#: 7





CERTIFICATE OF INSTALLATION

INSTRUMENT NAME : Rx Daytona +
SERIAL NUMBER : 7241-0663
CUSTOMER NAME : CluePath Diagnostic LLP
ADDRESS : # 19, New High School Road,
V.V Puram, Bangalore-560004

The undersigned performer certifies that the **Installation Qualification** protocol has been successfully completed for the instrument stated above.

Engineer

Signature 



Name: Samson Sundarraj

Designation: Customer Support Engineer

Company: Randox laboratories India Pvt Ltd.,



INSTALLATION QUALIFICATION

System Certification

Study data has been determined that the system described in this document either meets all criterial outlines in this Installation Qualification, or exceptional conditions have been identified and documentation included.

Exceptional conditions, if any have been addressed.

The system is ready for specified usage.

Protocol performed by: Randox Laboratories Representative
Name: Samson Sundarraj
Title: Customer Support Engineer
Company: Randox Laboratories India Pvt Ltd.

Customer authorization:

Name: Dr. MSN Prasad
Title: Sr. Pathologist & Managing Partner
CluePath Diagnostic
19, New High School Road, V V Puram
Bangalore 560004


Samson Sundarraj


Engineer Name & Signature
Date: 19 Jun 2023

Customer Signature
Date: 19 Jun 2023



Rx Daytona +

Calibration –Certificate –Rx Daytona +

Instrument	Rx Daytona + Clinical chemistry Analyzer
Serial No	7241-0663
Account Name	CluePath Diagnostic
Installation Date	19 Jun 2023
Calibration Date	19 Jun 2023
Next Calibration	18 Jun 2024

This is to certify that this analyzer has been inspected and calibration for following parameters

Test Parameter	Target value	Obtained value
Input Voltage	230-240 Volts	232 Volts
Incubator Temperature	37 ° C ± 0.3 Max	37 ° C ± 0.14
Reagent Tray Temperature	8-15 ° C	7.87 °C
Detector Performance Check All wavelengths	>8.3 Volts	~9.00 Volts
12V Lamp Supply	12 ± 0.3 volts	11.96 Volts
5V Supply	5 ± 0.3 volts	5.05 volts
24V Supply	24 ± 0.3 volts	24.03 volts
DTR Position	Between -20 & +20	Between -15 & +20
Water Blank	<5000	<1000

The results obtained are as per specifications & tolerance ranges. The above calibration was done with a Digital Multimeter (Model: MAS83OL, Serial Number: 404994) and Digital Thermometer (Model: Lutron/TM-917, Serial Number G.51771) Calibration certificate is enclosed.

For Randox Laboratories India Pvt Ltd


Authorized Signatory
Samson Sundarraj
Customer Support Engineer



