

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,  
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN: 33ADSPA9153Q1ZA • DL No: TN-12-20B -00398, TN-12-21B-00398

## CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Best Diagnostic Lab, #46A/1, Briyant Nagar 7 <sup>th</sup> Street, Near Town Bus stand , Thoothukudi.	Certificate No:	AMS/LAB/289/4417
		Issuing Date:	30/08/2023
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Micropipette	Id. No:	BD/LAB/VMP/001
Make:	Dragon Lab	Department:	Laboratory
Model:	---	Calibrated On:	28/08/2023
Sl. No:	YE213AS0153570	Calibration Due Suggested:	27/08/2024
Type:	Variable	DUC Condition on Receipt:	OK
Range & Least Count:	5 to 50µl		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/23/ATL/0512-02	31/01/2024
2	Thermo Hygrometer	Beetech/088	PC/01/23/ATL/0210-01	31/01/2024

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

### Test Results Volume µl:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	05	5.0	5.0	+0.0
2	10	10.0	10.0	+0.0
3	20	20.0	20.0	+0.0
4	30	30.0	30.0	+0.0
5	50	50.0	50.0	+0.0

Allowable Tolerance: User Defined

### Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test(i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

### Calibrated By:

(Calibration Engineer)



### Approved by:

(Technical Manager)

aceagm@gmail.com

amshealthcare@gmail.com

94434 74610, 93455 56148

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,  
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN: 33ADSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

## CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Best Diagnostic Lab, #46A/1, Briyant Nagar 7 <sup>th</sup> Street, Near Town Bus stand , Thoothukudi.	Certificate No:	AMS/LAB/289/4418
		Issuing Date:	30/08/2023
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Micropipette	Id. No:	BD/LAB/VMP/002
Make:	Dragon Lab	Department:	Laboratory
Model:	---	Calibrated On:	28/08/2023
Sl. No:	YE202AM0109017	Calibration Due Suggested:	27/08/2024
Type:	Variable	DUC Condition on Receipt:	OK
Range & Least Count:	10 to 100µl		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/23/ATL/0512-02	31/01/2024
2	Thermo Hygrometer	Beetech/088	PC/01/23/ATL/0210-02	31/01/2024

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

### Test Results Volume µl:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	10	10.0	10.0	+00.0
2	20	20.0	20.0	+00.0
3	50	50.0	50.0	+00.0
4	100	100.0	100.0	+00.0

Allowable Tolerance: User Defined

### Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test(i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

### Calibrated By:

(Calibration Engineer)

### Approved by:

(Technical Manager)



aceagm@gmail.com

amshealthcaremdu@gmail.com

94434 74610, 93455 56148



Scanned with OKEN Scanner



# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,  
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN. 33ADSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

## CALIBRATION CERTIFICATE

Customer Name & Address:	<b>M/S. Best Diagnostic Lab,</b> #46A/1, Briyant Nagar 7 <sup>th</sup> Street, Near Town Bus stand , Thoothukudi.	Certificate No:	AMS/LAB/289/4419
		Issuing Date:	30/08/2023
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	<b>Micropipette</b>	Id. No:	<b>BD/LAB/VMP/003</b>
Make:	Dragon Lab	Department:	Laboratory
Model:	---	Calibrated On:	<b>28/08/2023</b>
Sl. No:	YE226AX0393292	Calibration Due Suggested:	<b>27/08/2024</b>
Type:	Variable	DUC Condition on Receipt:	OK
Range & Least Count:	100 to 1000µl		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/23/ATL/0512-02	31/01/2024
2	Thermo Hygrometer	Beetech/088	PC/01/23/ATL/0210-02	31/01/2024

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

### Test Results Volume µl:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	100	100.0	100.0	+00.0
2	200	200.0	200.0	+00.0
3	500	500.0	500.0	+00.0
4	800	800.0	800.0	+00.0
5	1000	1000.0	1000.0	+00.0

Allowable Tolerance: User Defined

### Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test(i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

### Calibrated By:

(Calibration Engineer)

### Approved by:

(Technical Manager)

aceagm@gmail.com

amshealthcaremdu@gmail.com

94434 74610, 93455 56148

