

VIEW LAB MONTHLY SUMMARY

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Lab Name

Govt. PHC, Maraimalai Nagar

Lab No

10178

Month

September

Year

2023

Constituent Group

Chemistry III

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/09/2023

Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	318 mg/dL	0.48	0.69
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	187	35.19	12.73	4.48	40 mg/dL	1.07	0.66
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.1 mg/dL	-1.81	0.00
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	120 mg/dL	1.48	0.20
Interpretation										
SDI Range										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

MEDICAL OFFICER

Govt. Primary Health Center
Maraimalai Nagar- 603 209
Chengalpatu District