

VIEW LAB MONTHLY SUMMARY

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Lab Name **Govt. PHC, Maraimalai Nagar**  
 Lab No 10178  
 Month **October**  
 Year **2023**

Constituent Group **Chemistry III**

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : **20/10/2023**

Date of Report Published : 04/11/2023

| Sl. No.               | Analyte Name   | Analyzer Name                         | No. of Participants                      | CV     | CV    | SD   | CV        | CV    | CV   | CV |
|-----------------------|----------------|---------------------------------------|--|--------|-------|------|-----------|-------|------|----|
| 1                     | GLUCOSE III    | GOD-POD III                           | 1905                                     | 118.63 | 6.44  | 7.63 | 131 mg/dL | 1.62  | 0.35 |    |
| 2                     | UREA III       | Urease UV / GLDH III                  | 176                                      | 76.31  | 12.35 | 9.42 | 90 mg/dL  | 1.45  | 1.42 |    |
| 3                     | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | 1912                                     | 7.12   | 5.02  | 0.36 | 7.1 mg/dL | -0.06 | 0.02 |    |
| 4                     | CHOLESTROL III | CHOD-PAP III                          | 1786                                     | 111.59 | 4.28  | 4.78 | 111 mg/dL | -0.12 | 0.23 |    |
| <b>Interpretation</b> |                |                                       |  |        |       |      |           |       |      |    |
| <b>SDI Range</b>      |                |                                       |  |        |       |      |           |       |      |    |
| Within -1.00 to +1.00 |                |                                       | Excellent.                               |        |       |      |           |       |      |    |
| Within ±1.01 to ±2.00 |                |                                       | Good.                                    |        |       |      |           |       |      |    |
| Within ±2.01 to ±2.99 |                |                                       | Accept with caution. Warning Signal.     |        |       |      |           |       |      |    |
| Beyond ±3.0           |                |                                       | Unacceptable performance. Action Signal. |        |       |      |           |       |      |    |

**MEDICAL OFFICER**  
 Govt. Primary Health Center  
 Maraimalai Nagar - 603 209  
 Chengalpattu District