

Lab Name: Govt PSC, Mararamala Nagar

Lab No: 10178

Month: November

Year: 2023

Constituent Group: Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/11/2023

Date of Report Published : 05/12/2023

Sl. No	Analyte Name	Method	Value	Unit	SDI	Interpretation
1	GLUCOSE III	GOD-POD III	1894	mg/dL	5.36	Excellent
2	UREA III	Urease UV / GLDH III	187	mg/dL	12.39	Good
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	1871	mg/dL	6.15	Accept with caution Warning Signal
4	CHOLESTROL III	CHOD-PAP III	1778	mg/dL	5.29	Unacceptable performance Action Signal
SDI Range						
Within ±1.00 to ±1.00			Excellent			
Within ±1.01 to ±2.00			Good			
Within ±2.01 to ±2.99			Accept with caution Warning Signal			
Beyond ±3.0			Unacceptable performance Action Signal			



MEDICAL OFFICER

Govt Primary Health Center
Mararamala Nagar - 603 209
Cheng... natu District

Links