

CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - DECEMBER 2023



Lab Name: Govt. PHC, Maraimalai Nagar
Constituent Group: Chemistry III
PT Item: Lyophilized human serum based

Lab No: 10178
Date of Result Entered: 20/12/2023
Date of Report Published: 02/01/2024

Sl No	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	Participants	CV	Yield Value	SDI	U
1	GLUCOSE III	GOD-POD III	1885	99.99	6.05	6.05	6.05	101 mg/dL	0.17	0.28
2	UREA III	Urease UV / GLDH III	148	92.84	10.21	9.47	9.47	98 mg/dL	0.54	1.56
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	1844	1.74	6.08	0.11	0.11	1.63 mg/dL	-1.04	0.00
4	CHOLESTROL III	CHOD-PAP III	1761	107.76	4.90	5.28	5.28	115 mg/dL	1.37	0.25
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

LAB ADDRESS:
Govt. PHC, Maraimalai Nagar
MARAIMALAI NAGAR CLINICAL LABORATORY, CHENGULPET TALUK.,
Maraimalai Nagar
TAMILNADU603209

Coordinator Contact Details:
Email: cimgc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christodoss
Dr. Pamela Christodoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

MEDICAL OFFICER
Govt. Primary Health Center
Maraimalai Nagar- 603 209
Chengalpattu District