



CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/0412

Page No: 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,
MARAIMALAI NAGAR-600209, CHENGALPATTU DISTRICT.

SRF No.	SRF/23/00157-0020 & 09-02-2023
SRF Date	09-02-2023
Date of Receipt	09-02-2023
Date of Calibration	09-02-2023
Due Date for Calibration	08-02-2024
Issue Date	11/2/2023

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	UNIPRO
Range	5-50 μ l	Model	NA
Resolution	1 μ l	Material	PVC
Serial Number	LAB/MP/001	Operating Range	
ID Number		Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition			Calibration Method Used		
Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8635-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML05

Standard Used

Sl. No.	Description	ID.No. / Sl. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Semi Micro Balance	15112918	TVCSP/22/12/115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in μ l										
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (t)
		1	2	3	4	5				
1	10	9.93	9.91	9.92	9.91	9.93	9.92	-0.08	0.01	0.47
		9.95	9.90	9.91	9.91	9.92				
2	30	29.71	29.70	29.69	29.68	29.68	29.69	-0.31	0.02	0.47
		29.70	29.71	29.69	29.66	29.67				
3	50	49.10	49.11	49.13	49.10	49.11	49.11	-0.89	0.01	0.47
		49.14	49.11	49.11	49.10	49.09				

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00
- Calibration Liquid Used: Distilled or Deionized water (conductivity Grade 3) as specified in ISO 3696

Calibrated By:

Y. Ya.
(Calibration Engineer)
R. FAZLINYAN

Authorised by:

DM
(Quality Manager/Chief Executive)
D. VETRI SELVI

