



N.K Engineering & Research Center

Plot No. 142, Sneha Bharti Place, Mangal Chowk, Khemnichak, Ram Krishna Nagar, Patna, Bihar-800027

Mob: 8766317650, 7838663310

Website : www.nkerc.in , E-mail: info@nkerc.in , info.nkerc@gmail.com



CC-3349

Calibration Certificate

Page 1 of 1

Format No.-QS-22F

ULR No:- CC334923000003214F	Calibration Done on:- 26.09.2023	S.R.F. No-& Date:- 0220/23/09/2023
Issue Date:- 29.09.2023	Calibration Valid Upto:- 25.09.2024	Receipt Date:- 26.09.2023

Name & Address of Customer:-

M/s:- Biochemistry and Pathology Laboratory.

Regional Research Institute of Unani Medicine,

Guzri, Patna City, Bihar- 800008.

Identification of DUC (Device Under Calibration)

Instrument Name	Deep Freezer	Size/Range	-20 to 0 °C
Least Count	1 °C	Condition of Item	Good
Manufactured By	Optics Technology	Calibration Formed at	Site
Serial No.	5940	Location	Lab
Party ID Mark	5940	Ref. Standard	ITS-90

Calibration Procedure /Method: CI(T)-02

Environmental conditions: Temperature :25 ±3°C Relative Humidity: 50 ± 10 % RH

Standard & Major Equipment(s) Used:

Standard & Equipments	Range/ Serial No.	Calibration Agency	Traceability Cert. No.	Certificate Date	Due date
Digital Thermometer with 4 wire Rtd Sensor	(-80 to 400°C) / 20C107459/1053	Transcal	TSC/22-23/14983-2	10.01.2023	10.01.2024

Calibration Results

Measured Value on standar (in °C)	DUC Reading (in °C)	Deviation (in °C)	Related Uncertainty on C.F.(k=2) at approx 95% Confidence level ±
-15.007	-15	0.007	±0.6 °C
-10.005	-10	0.005	±0.6 °C
0.000	0	0.000	±0.6 °C

***** END OF REPORT *****

NOTE:

1. The result are valid only for this Calibrated item.
2. The calibration report shall not be reproduced without the written approval of calibration laboratory.
3. Result reported are valid at the time of & under the stated conditions of measurement
4. Laboratory standards are traceable to National/International Standards.
5. Calibration Certificates issued for scientific purpose only and should not be used for Trade / Commercial use.

CALIBRATED BY


(CALIBRATION ENGINEER)



APPROVED BY


NIRHIL KUMAR JHA
(Q.M.T.M.)