



N.K Engineering & Research Center

Plot No. 142, Sneha Bharti Place, Mangal Chowk, Khemnichak, Ram Krishna Nagar, Patna, Bihar-800027

Mob: 8766317650, 7838663310

Website : www.nkerc.in , E-mail: info@nkerc.in, info.nkerc@gmail.com

Calibration Certificate

Page NO. :- 1 of 1

Formate No:- QS-22F

Certificate No-NKERC/23-24/M/S/26.11	Calibration Done on:- 26.09.2023	S.R.F.No-& Date:- 0220/23/09/2023
Issue Date:- 29.09.2023	Calibration Valid Upto:- 25.09.2024	Receipt Date:- 26.09.2023
Name & Address of Customer:- M/s:- Biochemistry and Pathology Laboratory. Regional Research Institute of Unani Medicine, Guzri, Patna City, Bihar- 800008.		

Identification of UUC (Unit Under Calibration)

Instrument Name	Centrifuge	Size/Range	5280 RPM
Least Count	10 RPM	Condition of Item	Good
Manufactured By	REMI/R-8C	Calibration Performed at	Site
Manufacturer M/Sr. No.	AHLC-1202	Location	Lab/REMI TABLE TOP
Party ID Mark	AHLC-1202	Ref. Standard	Is 12508

Environmental Condition (s): Temperature : 25± 2 °C. Relative Humidity :55 ± 10 % RH

Standard & Major Equipment (s) Used :

Standard(s) & Equipments	Range/ Serial No.	Calibrating Agency	Traceability Cert. No.	Certificate Date	Valid Up to
Digital Techometer	5 TO 99999RPM/ L182537	NKERC/23-24/ M/S/17.04	CC23122200000 4636F	17.01.2023	16.01.2024

Uncertainty of measurement (at Approx 95% Confidence Level With Coverage Factor k = 2) = ± 0.04 %

Calibration Results:

UUC Reading (RPM)	Master Reading (RPM)	Error (RPM)
1000	1000.23	-0.23
2000	2000.32	-0.32
3000	3000.41	-0.41
4000	4000.87	-0.87
5000	5000.93	-0.93

***** END OF REPORT *****

NOTE:

1. The result are valid only for this Calibrated item.
2. The calibration report shall not be reproduced without the written approval of calibration laboratory.
3. Result reported are valid at the time of & under the stated conditions of measurement
4. Laboratory standards are traceable to National/International Standards.
5. Calibration Certificates issued for scientific purpose only and should not be used for Trade / Commercial use.

CALIBRATED BY


(CALIBRATION ENGINEER)



APPROVED BY


NIKHIL KUMAR JHA
(Q.M/T.M.)