



N.K Engineering & Research Center

Plot No. 142, Sneha Bharti Place, Mangal Chowk, Khemnichak, Ram Krishna Nagar, Patna, Bihar-800027
Mob: 8766317650, 7838663310

Website : www.nkerc.in , E-mail: info@nkerc.in , info.nkerc@gmail.com



Calibration Certificate

Page 1 of 1

Format No.-QS-22F

ULR No:- CC334923000003211F	Calibration Done on:- 26.09.2023	S.R.F. No-& Date:- 0220/23/09/2023
Issue Date:- 29.09.2023	Calibration Valid Upto:- 25.09.2024	Receipt Date:- 26.09.2023

Name & Address of Customer:-

M/s:- Biochemistry and Pathology Laboratory.
Regional Research Institute of Unani Medicine,
Guzri, Patna City, Bihar- 800008.

Identification of DUC (Device Under Calibration)

Instrument Name	Hot Air Oven	Size/Range	0 to 250 °C
Least Count	1 °C	Condition of Item	Good
Manufactured By	Optics Technology	Calibration Formed at	Site
Serial No.	14EQO2	Location	Lab
Party ID Mark	14EQO2	Ref. Standard	ITS-90

Calibration Procedure /Method: CI(T)-02

Environmental conditions: Temperature :25 ±3°C Relative Humidity: 50 ± 10 % RH

Standard & Major Equipment(s) Used:

Standard & Equipments	Range/ Serial No.	Calibration Agency	Traceability Cert. No.	Certificate Date	Due date
Digital Thermometer with 4 wire Rtd Sensor	(-80 to 400°C)/ 20C107459/1053	Transcal	TSC/22-23/14983-2	10.01.2023	10.01.2024

Calibration Results

Measured Value on standar (in °C)	DUC Reading (in °C)	Deviation (in °C)	Related Uncertainty on C.F.(k=2) at approx 95% Confidence level ±
49.941	50	0.059	±0.6 °C
99.888	100	0.112	±0.6 °C
149.889	150	0.111	±0.7 °C
199.801	200	0.199	±0.7 °C
249.759	250	0.241	±0.7 °C

***** END OF REPORT *****

NOTE:

- The result are valid only for this Calibrated item.
- The calibration report shall not be reproduced without the written approval of calibration laboratory.
- Result reported are valid at the time of & under the stated conditions of measurement
- Laboratory standards are traceable to National/International Standards.
- Calibration Certificates issued for scientific purpose only and should not be used for Trade / Commercial use.

CALIBRATED BY

(CALIBRATION ENGINEER)



APPROVED BY

NIKHIL KUMAR JHA
(Q.M/T.M.)