N.K Engineering & Research Center

Plot No. 142, Sneh Bharti Place, Mangal Chowk, Khemnichak, Ram Krishna Nagar, Patna, Bihar-800027
Mob: 8766317650, 7838663310
Website : www.nkerc.in , E-mail: info@nkerc.in , info.nkerc@gmail.com

Calibration Certificate

Certificate No.:- NKERC/23-24/M/S/26.08	Calibration Done on:- 26.09.2023	S.R.F. No-& Date:- 0220/23/09/2023
Issue Date:- 29.09.2023	Calibration Valid Upto:- 25.09.2024	Receipt Date:- 26.09.2023

Name & Address of Customer:-

M/s:- Biochemistry and Pathology Laboratory.

Regional Research Institute of Unani Medicine,

Guzri, Patna City, Bihar- 800008.

Identification of UUC (Unit Under Calibration)

Instrument Name	Micropipette	Size/Range	100 to 1000 µl
Least Count	5 µl	Condition of Item	Good
Manufactured By	ERBA	Calibration Formed at	Lab
Serial No.	MB400410	Location	Lab
Party ID Mark	MB400410	Ref. Standard	ISO-4787

Calibration Procedure /Method: CI(T)-02

Environmental conditions: Temperature :25 ±3°C Relative Humidity: 50 ± 10 % RH

Standard & Major Equipment(s) Used:

Standard & Equipments	Range/ Serial No.	Calibration Agency	Tracebility Cert. No.	Certificate Date	Due date
Digital Weighing Balance	1 mg to 200 g	Tycon	TYCON/W/09/2023/860	13.09.2023	12.09.2024

Calibration Results

| UUC Value (µl)	Observed Value (µl)	Error (µl)
100	99.991	0.009
200	199.988	0.012
400	399.985	0.015
600	599.980	0.020
800	799.986	0.014
1000	999.972	0.028

**********END THE REPORT**********

NOTE:
1. The result are valid only for this Calibrated item.
2. The calibration report shall not be reproduced without the written approval of calibration laboratory.
3. Result reported are valid at the time of & under the stated conditions of measurement
4. Laboratory standards are traceable to National/International Standards.
5. Calibration Certificates issued for scientific purpose only and should not be used for Trade / Commercial use.

CALIBRATED BY

(CALIBRATION ENGINEER)

APPROVED BY

NIKHIL KUMAR JHA
(Q.M/T.M.)