

N.K Engineering & Research Center

Plot No. 142, Sneh Bharti Place, Mangal Chowk, Khemnichak, Ram Krishna Nagar, Patna, Bihar-800027 Mob: 8766317650, 7838663310

Website: www.nkerc.in, E-mail: info@nkerc.in, info.nkerc@gmail.com

Calibration Certificate

Page 1 of 1

Format No.-QS-22F

Certificate No.:- NKERC/23-24/M/S/26.06	Calibration Done on:- 26.09.2023	S.R.F. No-& Date:- 0220/23/09/2023	
Issue Date:- 29.09.2023	Calibration Valid Upto:- 25.09.2024	Receipt Date:- 26.09.2023	

Name & Address of Customer:-

M/s:- Biochemistry and Pathology Laboratory.

Regional Research Institute of Unani Medicine,
Guzri, Patna City, Bihar- 800008.

Instrument Name	Micropinetto	S: /D	0.1-100-1
instrument Name	Micropipette	Size/Range	0 to 100 μl
Least Count	1 μΙ	Condition of Item	Good
Manufactured By	Perfect	Calibration Formed at	Lab
Serial No.	13PO1	Location	Lab
Party ID Mark	13PO1	Ref. Standard	ISO-4787

Calibration Procedure / Method: CI(T)-02

Environmental conditions: Temperature :25 \pm 3°C Relative Humidity: 50 \pm 10 % RH

Standard & Major Equipment(s) Used:

Standard & Equipments	Range/ Serial No.	Calibration Agency	Tracebility Cert. No.	Certificate Date	Due date
Digital Weighing Balance	1 mg to 200 g	Tycon	TYCON/W/09/ 2023/860	13.09.2023	12.09.2024

Calibration Results

UUC Value	Observed Value	Error	
(μl)	(μΙ)	(µI)	
10	10 10.007		
20	20.011	-0.011	
40	40.017	-0.017	
60	60.021	-0.021	
80	80.027	-0.027	
100	100.032	-0.032	

*********END THE REPORT*******

NOTE

- 1. The result are valid only for this Calibrated item.
- 2. The calibration report shall not be reproduced without the written approval of calibration laboratory.
- 3. Result reported are valid at the time of & under the stated conditions of measurement
- Laboratory standards are traceable to National/International Standards.
- 5. Calibration Certificates issued for scientific purpose only and should not be used for Trade / Commercial use.

CALIBRATED BY

(CALIBRATION HNGINEER)



APPROVED BY

HIL KUMAR (O.M/T.M.)