



ADL/KO/SER/0335/23-24

CERTIFICATE OF CALIBRATION

Customer Name : Precise Health Care

Address of the customer : 27/1, 4th Main, Dattareyanagar, Hosakarehalli, Bangalore

Date of Calibration : 25/JAN/2024

Calibration Due on : 24/JUL/2024

Instrument Model : BC 3000 Plus – 3 Part Hematology Analyzer

Serial No. : RJ-9C129347

Date of Installation : 11/APR/2014

This is to certify that the above-mentioned product has been calibrated for CBC parameters (WBC, RBC, HGB, MCV, PLT) according to the standard operating procedure given by the MINDRAY.

The reference instruments used for value assignment are managed by the traceability system in Shenzhen Mindray Bio – Medical Electronics Co., Ltd, and these are traceable to the International Standards such as ICSH/WHO, NCCLS.

For AGAPPE DIAGNOSTICS LTD.



Authorized Signatory

Name: Nijin C

Designation: Engineer - Customer Service

AGAPPE DIAGNOSTICS LTD.

ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
406, Merfin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in

MUMBAI (REGISTERED OFFICE)
401 & 402, 4th Floor, Jaishingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
5-6, IInd Floor, Red Cross Bhavan, No:26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in

**DETAILS OF CONTROL RUN**

Instrument Model : BC 3000 PLUS - 3 Part Hematology Analyzer
 Serial No : RJ-9C129347
 Date of Control Run : 25/01/2024
 Control Used : HEMATOLOGY TRI LEVEL CONTROL
 Lot No : 32310038
 Expiry : 05-03-2024

The results of the repeatability test are summarized in the following table. Refer the attached control sheets for details.

Sample No	WBC K/ μ L	RBC M/ μ L	HGB g/dL	MCV fL	PLT K/ μ L
1	8.1	4.31	13	82.9	247
2	7.9	4.38	13	82.9	252
3	8.2	4.36	13.2	82.9	245
4	7.7	4.40	13.1	82.9	241
Instrument Mean	7.8	4.2	12.8	84.1	231
CV %	2.7	0.08	0.73	0	1.8
Factor	101.1	103.4	100.9	99.8	106.3

This is to certify that the above-mentioned product has been calibrated for CBC parameters (WBC, RBC, HGB, MCV, PLT) according to the standard operating procedure given by the Mindray.

The control run after calibration shows good repeatability and precision.

Next Calibration due on **24/JUL/2024**

For AGAPPE DIAGNOSTICS LTD.

For



Authorized Signatory

Name: Nijin C

Designation: Engineer - Customer Service

AGAPPE DIAGNOSTICS LTD.

ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
 Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
 Tel: +91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
 X/588-CB, Block No. 32, KINFRA Small Industrial Park,
 Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
 406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
 Salt Lake City, Kolkata - 700 091.
 Tel: +91 33 4003 0451 | Email: kolkattooffice@agappe.in

MUMBAI (REGISTERED OFFICE)
 401 & 402, 4th Floor, Jaisingh Business Centre, 119,
 Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
 Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
 DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
 New Delhi - 110 015, India.
 Tel: +91 11 4558 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
 5-6, 11th Floor, Red Cross Bhavan, No-26 Race Course Road,
 Bangalore - 560 001. Tel: +91 80 2228 8288
 Email: bangaloreoffice@agappe.in

**CALIBRATION DATA**

CHANNEL	FACTOR	DATE
WBC	101.1	25-01-2024
RBC	103.4	25-01-2024
HGB	100.9	25-01-2024
MCV	99.8	25-01-2024
PLT	106.3	25-01-2024

For AGAPPE DIAGNOSTICS LTD.

For 
Authorized Signatory

Name: Nijin C

Designation: Engineer - Customer Service

**AGAPPE DIAGNOSTICS LTD.** ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
Tel: + 91 484 286 7000 | Email: agappe@agappe.in

MUMBAI (REGISTERED OFFICE)
401 & 402, 4th Floor, Jaisingh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbai@agappe.in

EQUIPMENT PLANT
X/588-CB, Block No. 32, KINFRA Small Industrial Park,
Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

DELHI OFFICE
DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4558 8416 | Email: delhi@agappe.in

KOLKATA OFFICE
406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkata@agappe.in

BANGALORE OFFICE
5-6, 11th Floor, Red Cross Bhavan, No:26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangalore@agappe.in

Manual

Ready

Whole



18:23

Para	Factor	Time
WBC	101.1 %	25-01-2024
RBC	103.4 %	25-01-2024
HGB	100.9 %	25-01-2024
MCV	99.8 %	25-01-2024
PLT	106.3 %	25-01-2024