



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



Govt. PHC, Malamanjanur

Lab No

12935

ent Group

Chemistry III

Date of Result Entered :

22/08/2023

Lyophilized human serum based

Date of Report Published :

01/09/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	176 mg/dL	0.05	0.42
UREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.7 mg/dL	-0.50	0.01
CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	112 mg/dL	-0.39	0.22

SDI Range	Interpretation
-1.00 to +1.00	Excellent.
-2.01 to ±2.00	Good.
-3.01 to ±2.99	Accept with caution. Warning Signal.
±3.0	Unacceptable performance. Action Signal.

ADDRESS :

Govt. PHC, Malamanjanur

Malamanjanur

Vellore, Tamil Nadu 606708

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

மருத்துவ அலுவலர்,
அரசு ஆரம்ப சுகாதார நிலையம்,
மலமஞ்சனூர் - 606 708,
வெ.நா. கி.மலை மாவட்டம்.



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Govt. PHC, Malamanjanur

Lab No

12935

Chemistry III

Date of Result Entered :

20/09/2023

Lyophilized human serum based

Date of Report Published :

05/10/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	315 mg/dL	0.28	0.69
UREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.4 mg/dL	1.05	0.00
CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	114 mg/dL	0.09	0.20

SDI Range	Interpretation
-0.00 to +1.00	Excellent.
-0.01 to ±2.00	Good.
-0.01 to ±2.99	Accept with caution. Warning Signal.
>±3.0	Unacceptable performance. Action Signal.

ADDRESS :
Govt. PHC, Malamanjanur
Malamanjanur
DU606708

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

மருத்துவ அலுவலர்
அரசு ஆரம்ப சுகாதார நிலையம்
மலமஞ்சனூர் - 606 708
கோட்டாம்பட்டி வட்டம், தி.மலை மாவட்டம்



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - OCTOBER 2023



PC-1024

ne Govt. PHC, Malamanjanur Lab No 12935
ent Group Chemistry III Date of Result Entered : 20/10/2023
Lyophilized human serum based Date of Report Published : 04/11/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1905	118.63	6.44	7.63	110 mg/dL	-1.13	0.35
CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1912	7.12	5.02	0.36	7.5 mg/dL	1.06	0.02
CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1786	111.59	4.28	4.78	104 mg/dL	-1.59	0.23

SDI Range	Interpretation
1.00 to +1.00	Excellent.
1.01 to ±2.00	Good.
2.01 to ±2.99	Accept with caution. Warning Signal.
±3.0	Unacceptable performance. Action Signal.

ADDRESS :
Govt. PHC, Malamanjanur
Malamanjanur
Vellore
TADU606708

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not subcontract any components
***** End of Report *****

Pamela Christudoss
மருத்துவ அலுவலர்,
அரசு ஆரம்ப சுகாதார நிலையம்,
மலமஞ்சனூர் - 606 708.
ண்டராம்பட்டு வட்டம், கி.மலை மாவட்டம்



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - NOVEMBER 2023



PC-1024

Govt. PHC, Malamanjanur
Chemistry III
Lyophilized human serum based

Lab No 12935
Date of Result Entered : 20/11/2023
Date of Report Published : 05/12/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1894	218.58	5.36	11.73	225 mg/dL	0.55	0.54
UREA NITROGEN III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1871	4.36	6.15	0.27	4.2 mg/dL	-0.60	0.01
CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1778	106.36	5.29	5.62	109 mg/dL	0.47	0.27

SDI Range	Interpretation
0.00 to +1.00	Excellent.
0.01 to ±2.00	Good.
0.01 to ±2.99	Accept with caution. Warning Signal.
3.0	Unacceptable performance. Action Signal.

ADDRESS :
Malamanjanur
anur
U606708

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

மருத்துவ அலுவலர்
அரசு ஆரம்ப சுகாதார நிலை
மலமஞ்சளூர் - 606 708.
ஸ்ட்ராம்பட்டு லப.ம். சி.ம.க.வெல்லூர்



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - DECEMBER 2023



PC-1024

Client Name: Govt. PHC, Malamanjanur
Client Group: Chemistry III
Lyophilized human serum based
Lab No: 12935
Date of Result Entered: 20/12/2023
Date of Report Published: 02/01/2024

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1885	99.99	6.05	6.05	99 mg/dL	-0.16	0.28
CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1844	1.74	6.08	0.11	2 mg/dL	2.45	0.00
CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	107.76	4.90	5.28	106 mg/dL	-0.33	0.25

SDI Range	Interpretation
1.00 to +1.00	Excellent.
1.01 to ±2.00	Good.
2.01 to ±2.99	Accept with caution. Warning Signal.
±3.0	Unacceptable performance. Action Signal.

ADDRESS :
C, Malamanjanur
njanur
DU606708

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

Pamela Christudoss
மருத்துவ அலுவலர்.
அரசு ஆரம்ப சுகாதார நிலையம்.
மலமஞ்சனூர் - 606 708.
ஊடகமட்டு வட்டம், சி.மலை மாவட்டம்



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JANUARY 2024



PC-1024

Govt. PHC, Malamanjanur
Chemistry III
Lyophilized human serum based

Lab No 12935
Date of Result Entered : 20/01/2024
Date of Report Published : 06/02/2024

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1882	145.08	6.30	9.14	157 mg/dL	1.30	0.42
CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1872	1.64	7.16	0.12	1.6 mg/dL	-0.34	0.01
CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	153	107.91	6.24	6.74	106 mg/dL	-0.28	1.09

SDI Range	Interpretation
1.00 to +1.00	Excellent.
1.01 to ±2.00	Good.
2.01 to ±2.99	Accept with caution. Warning Signal.
±3.0	Unacceptable performance. Action Signal.

ADDRESS :
Govt. PHC, Malamanjanur
Malamanjanur
Vellore
TADU606708

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

Pamela Christudoss
மருத்துவ அலுவலர்
அரசு ஆரம்ப சுகாதார நிலை
மலமஞ்சனூர் - 606 708.
ண்டராம்பட்டு வட்டம், தி.மலை மாவட்டம்