

Instrument : Abbott ARCHITECT ci4100

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Urea	mg/dl	21.00	22.03	0.24	0.25
@ Creatinine	mg/dl	0.71	0.74	-0.20	0.07
@ Bilirubin Total	mg/dl	0.20	0.24	-0.06	0.12
@ Total Protein	g/dL	6.00	5.80	0.97	0.80
@ Albumin	g/dL	3.30	3.24	0.03	0.06
@ Uric Acid	mg/dl	3.80	3.68	0.18	0.35
@ Calcium	mg/dl	6.90	7.12	-0.53	-0.70
@ Phosphorous Inorganic	mg/dl	3.70	3.68	-0.11	-0.03
@ Cholesterol	mg/dl	130.00	129.67	-0.07	0.60
@ Triglycerides	mg/dl	127.00	125.00	-0.28	-0.55
@ HDL	mg/dl	33.00	32.83	0.33	0.56
@ LDL	mg/dl	68.00	67.87	-0.67	-0.48
@ ALT	U/L	14.00	13.17	0.47	0.30
@ AST	U/L	21.00	23.75	-1.02	-0.69
@ ALP	U/L	44.00	46.67	-0.15	-0.27
@ GGT	U/L	27.00	26.50	0.38	0.68
@ Amylase	U/L	56.00	56.33	0.00	0.00

Instrument : Erba EC 90

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	77.90	79.03	0.02	0.66
@ Sodium	mmol/L	103.00	104.37	-0.71	0.30
@ Potassium	mmol/L	3.34	3.44	-0.55	-0.32

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning

X : z score ≥ 3 - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	20
Not Evaluated Parameters	0
Evaluated Parameters	20
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirod Road, Kurla (W),
Mumbai - 400070

Instrument : Abbott ARCHITECT ci4100

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	70.00	70.00	-0.80	-0.66
@ Urea	mg/dl	20.00	21.03	0.25	0.25
@ Creatinine	mg/dl	1.60	1.62	-0.08	0.07
@ Bilirubin Total	mg/dl	2.60	2.81	0.29	0.12
@ Bilirubin Direct	mg/dl	0.40	0.39	0.43	0.43
@ Total Protein	g/dL	5.90	5.83	0.62	0.80
@ Albumin	g/dL	3.40	3.38	-0.01	0.06
@ Uric Acid	mg/dl	3.90	3.80	0.26	0.35
@ Calcium	mg/dl	7.00	7.02	-0.36	-0.70
@ Phosphorous Inorganic	mg/dl	3.30	3.20	0.05	-0.03
@ Cholesterol	mg/dl	144.00	134.33	1.26	0.60
@ Triglycerides	mg/dl	113.00	112.33	-0.81	-0.55
@ HDL	mg/dl	21.00	20.67	0.78	0.56
@ LDL	mg/dl	80.00	82.27	-0.28	-0.48
@ ALP	U/L	41.00	46.83	-0.39	-0.27
@ GGT	U/L	27.00	24.00	0.97	0.68
@ Amylase	U/L	54.00	54.00	0.00	0.00

Instrument : Erba EC 90

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	81.20	81.51	0.44	0.66
@ Sodium	mmol/L	103.40	107.08	-1.30	0.30
@ Potassium	mmol/L	3.41	3.57	-0.60	-0.32

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning

X : z score ≥ 3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	20
Not Evaluated Parameters	0
Evaluated Parameters	20
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070

Instrument : Abbott ARCHITECT ci4100

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	73.00	73.80	-0.51	-0.66
@ Creatinine	mg/dl	0.75	0.72	0.49	0.07
@ Albumin	g/dL	3.90	3.85	0.15	0.06
@ Uric Acid	mg/dl	4.30	4.14	0.62	0.35
@ Calcium	mg/dl	7.50	7.68	-1.21	-0.70
@ ALT	U/L	6.00	5.60	0.13	0.30
@ AST	U/L	13.00	14.00	-0.35	-0.69

Instrument : Erba EC 90

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	112.70	116.22	1.53	0.66
@ Sodium	mmol/L	183.80	183.79	0.00	0.30
@ Potassium	mmol/L	7.00	7.02	0.19	-0.32

Legend @ : Acceptable
 † : 2.0 < z score < 3.0 - Warning
 X : z score ≥3 - Unacceptable
 # : Not Evaluated
 ⊖ : Delayed Result Entry
 * : Not considered for evaluation.

Total Parameters	10
Not Evaluated Parameters	0
Evaluated Parameters	10
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirol Road, Kurla (W),
 Mumbai - 400070