

Instrument : Horiba Yumizen H550

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	0.99	1.02	-0.30	-0.21
@ Total Leucocytes (WBC) count	/c.mm	1680.00	1835.53	-0.54	-0.44
@ Haemoglobin	g/dL	2.70	2.68	0.00	0.05
@ MCV (Mean Corpuscular Volume)	fL	129.20	125.66	0.23	0.21
@ MCH (Mean Corpuscular Hb)	pg	27.00	26.65	0.39	0.29
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	20.90	20.97	0.16	0.07
@ Platelet Count	1000/ $\mu$ L	75.00	87.19	-0.46	-0.45
@ Hematocrit	%	12.80	12.66	-0.21	-0.09
@ RDW CV%	%	26.10	24.07	0.63	1.15

Legend @ : Acceptable

! :  $2.0 < z \text{ score} < 3.0$  - Warning


X :  $z \text{ score} \geq 3$  - Unacceptable

# : Not Evaluated

⌚ : Delayed Result Entry

\* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



**Dr Puneet Kumar Nigam**  
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070

Instrument : Horiba Yumizen H550

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	1.47	1.45	-0.11	-0.21
@ Total Leucocytes (WBC) count	/c.mm	3300.00	3456.41	-0.33	-0.44
@ Haemoglobin	g/dL	4.10	4.05	0.10	0.05
@ MCV (Mean Corpuscular Volume)	fL	127.30	124.09	0.19	0.21
@ MCH (Mean Corpuscular Hb)	pg	27.90	27.88	0.19	0.29
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	21.90	22.18	-0.02	0.07
@ Platelet Count	1000/ $\mu$ L	134.00	151.44	-0.43	-0.45
@ Hematocrit	%	18.70	17.87	0.04	-0.09
@ RDW CV%	%	27.20	22.46	1.67	1.15

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score  $\geq$ 3 - Unacceptable

# : Not Evaluated

⌚ : Delayed Result Entry

\* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



**Dr Puneet Kumar Nigam**  
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070