


Instrument : Abbott ARCHITECT ci4100

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ HbA1c	%	6.40	6.93	-1.84	-1.56

Legend @ : Acceptable	Total Parameters	1
! : 2.0 < z score < 3.0 - Warning	Not Evaluated Parameters	0
X : z score ≥3 - Unacceptable	Evaluated Parameters	1
# : Not Evaluated	Outlier Parameters (X)	0
⌚ : Delayed Result Entry	EQAS Score HbA1c	100.00 %
* : Not considered for evaluation.		



**Dr Puneet Kumar Nigam**  
PT coordinator & Technical Manager, MHL EQAS  
Unit No. 409-416.  
Commercial Building - 1A  
Kohinoor Mall, Kirol Road, Kurla (W),  
Mumbai - 400070

Instrument : Abbott ARCHITECT ci4100

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ HbA1c	%	5.60	5.77	-1.28	-1.56

Legend @ : Acceptable	Total Parameters	1
! : 2.0 < z score < 3.0 - Warning	Not Evaluated Parameters	0
X : z score ≥3 - Unacceptable	Evaluated Parameters	1
# : Not Evaluated	Outlier Parameters (X)	0
⌚ : Delayed Result Entry	EQAS Score HbA1c	100.00 %
* : Not considered for evaluation.		



**Dr Puneet Kumar Nigam**  
PT coordinator & Technical Manager, MHL EQAS  
Unit No. 409-416.  
Commercial Building - 1A  
Kohinoor Mall, Kiroil Road, Kurla (W),  
Mumbai - 400070