

Erba Lachema s.r.o.



Brno, 10.11.2016

To,  
Redcliffe Lifetech Private Limited.  
Majura Gate , Surat, Gujrat.

Instrument Name: ERBA LAURA

Instrument Serial No: 1412-1-  
202305

We Erba Lachema SRO with registered office in Czech Republic, Karasek 2219/1D, 621 00 BMO, represented by its Production manager Veronika Pleskova here by confirm as Producer of LAURA instrument that this device is firmly calibrated by the manufacturer and does not require any regular calibration by the user, correct calibration of instrument is controlled by grey strips provided initially together with the instrument.

*P. Suchan*

RNDr. Pavel Suchan Ph.D.

Global Product Manager, Urine Analysis

Erba Lachema s.r.o.

Karásek 1d, Brno 62100, Czech Republic





Urinalysis

LAURA

\*\*\*\*\*

2024.02.22

11:45 AM

DEKAPHAN LAURA

Seq.No: 0001

URINORM N

2024.02.22

11:49 AM

BLD	NEG
BIL	NEG
UBG	NORM
KET	NEG
GLU	NEG
PRO	NEG
pH	6
NIT	NEG
LEU	NEG
SG	1.020



QC TEST: OK

DEKAPHAN LAURA

Seq.No: 0002

URINORM P

2024.02.22

11:57 AM

*BLD	10	Ery/u1
*BIL	3	mg/dl
*UBG	1	mg/dl
*KET	52	mg/dl
*GLU	500	mg/dl
*PRO	500	mg/dl
pH	9	
*NIT	POS	
*LEU	500	Leu/u1
SG	1.000	



QC TEST: OK