


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 240123A/N-110	Page	: 1 of 1
CALIBRATED ON.	: 25 / 01 / 2024	Ref. Std.	: IS:12508:1988
CALIBRATION DUE.	: 24 / 01 / 2025	C.P. No.	: ACS/CP/10
SRF No. & DATE	: 240123A/N Dt. 23.01.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC34272400000119F Dt. 27.01.2024	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. Ground Floor Shop No 1, Leela Health Care, Kailash Nagar, Majura Gate, Surat, Gujrat 395001		

ITEM CALIBRATED	: Centrifuge		
Make	: Remi	Range	: (0 to 6000) RPM
L. Count	:	ID No.	: RCL/BIO/Cent-01
Location	: Clinical Biochemistry	S. No.	: ZIKN-43654

ENVIRONMENT CONDITION: Temp. { 30°C ± 10°C } R.H. {30% to 80%}

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	G-230908-1-1	08/09/2024

RESULTS : Mechanical Discipline			
* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)	
1000	1001.3	- 1.3	
2000	2002.5	- 2.5	
4000	4004.8	- 4.8	
6000	6006.6	- 6.6	
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 3.64 rpm of reading at coverage Factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by *[Signature]*

....End of Calibration Certificate....

