


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 240123A/N-116	Page	: 1 of 1
CALIBRATED ON.	: 25 / 01 / 2024	Ref. Std.	: CG-11
CALIBRATION DUE.	: 24 / 01 / 2025	C.P. No.	: ACS/CP/07
SRF No. & DATE	: 240123A/N Dt. 23.01.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342724000000125F Dt. 27.01.2024	Cal. At	: Site / Lab

CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. Ground Floor Shop No 1, Leela Health Care, Kailash Nagar, Majura Gate, Surat, Gujrat 395001
-----------------------	--

ITEM CALIBRATED	: Digital Temp. Indicator
Make	: Range : (-50 to 110) °C
L. Count	: 0.1°C ID No. : RCL/BIO/D.TEMP-01
Location	: Clinical Biochemistry S. No. : N.A

ENVIRONMENT CONDITION:	Temp. { 30°C ± 10°C } R.H. {30% to 80%}
-------------------------------	---


STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Universal Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/041123/1-1	07/11/2024

RESULTS : ET Discipline			
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)	
2.2	2	0.2	
4.3	4	0.3	
6.4	6	0.4	
8.5	8	0.5	
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by 

....End of Calibration Certificate....

