


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)  
Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b> : 240123A/N-117		<b>Page</b> : 1 of 1		
<b>CALIBRATED ON.</b> : 25 / 01 / 2024		<b>Ref. Std.</b> : CG-11		
<b>CALIBRATION DUE.</b> : 24 / 01 / 2025		<b>C.P. No.</b> : ACS/CP/07		
<b>SRF No. &amp; DATE</b> : 240123A/N Dt. 23.01.2024		<b>Phy. Inspection</b> : O.K.		
<b>ULR No. &amp; ISSUE DATE</b> : CC342724000000126F Dt. 27.01.2024		<b>Cal. At</b> : Site / Lab		
<b>CALIBRATED FOR</b> : M/s Redcliffe Lifetech Pvt. Ltd. Ground Floor Shop No 1, Leela Health Care, Kailash Nagar, Majura Gate, Surat, Gujrat 395001				
<b>ITEM CALIBRATED</b> : Digital Temp. Indicator				
Make :		Range : (-50 to 110) °C		
L. Count : 0.1°C		ID No. : RCL/BIO/D.TEMP-02		
Location : Clinical Biochemistry		S. No. : N.A		
<b>ENVIRONMENT CONDITION:</b> Temp. { 30°C ± 10°C} R.H. {30% to 80%}				
<b>STANDARD INSTRUMENTS USED :</b>				
<b>INSTT. NAME</b>	<b>ID No. / S. No.</b>	<b>TRACEABILITY</b>	<b>CERTIFICATE No.</b>	<b>DUE DATE</b>
Universal Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/041123/1-1	07/11/2024
<b>RESULTS</b> : ET Discipline				
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)		
2.0	2	-		
4.1	4	0.1		
6.2	6	0.2		
8.3	8	0.3		
(average of readings)				

### REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=2.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmary statement not provided since not required.

Prepared by *Pei*

....End of Calibration Certificate....

