


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | |
|---------------------------------|-------------------------------------|------------------------|--------------|
| CERTIFICATE NO. | : 240123A/N-102 | Page | : 1 of 1 |
| CALIBRATED ON. | : 23 / 01 / 2024 | Ref. Std. | : ISO:8655:6 |
| CALIBRATION DUE. | : 22 / 01 / 2025 | C.P. No. | : ACS/CP/08 |
| SRF No. & DATE | : 240123A/N Dt. 23.01.2024 | Phy. Inspection | : O.K. |
| ULR No. & ISSUE DATE | : CC342724000000111F Dt. 27.01.2024 | Cal. At | : Site / Lab |

| | |
|-----------------------|--|
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. Ground Floor Shop No 1, Leela Health Care, Kailash Nagar, Majura Gate, Surat, Gujrat 395001 |
|-----------------------|--|

| | |
|------------------------|---------------------|
| ITEM CALIBRATED | : Micro Pipette |
| Make | : Thermo Scientific |
| Range | : (10 to 100) µl |
| L. Count | : |
| ID No. | : RCL/HEM/PIP-02 |
| Location | : Hematology |
| S. No. | : TH48644 |

| | |
|-------------------------------|--|
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 4°C } R.H. { 30% to 75% } |
|-------------------------------|--|

| | | | | |
|------------------------------------|------------------------|---------------------|------------------------|-----------------|
| STANDARD INSTRUMENTS USED : | | | | |
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Weighing Balance | ACS/DB/02/D449927064 | BlueBox | BBT/691/JAN/24 | 04/01/2025 |

| | | | | |
|------------------------------|------------------------|---------------------------|-----------------------------|-------------------|
| RESULTS : | | | | |
| Mechanical Discipline | | | | |
| Nominal Capacity (µl) | Graduation (µl) | Nominal Value (µl) | Volume at 20 °C (µl) | Error (µl) |
| | | 10 | 9.54 | 0.46 |
| 100 | | 50 | 49.75 | 0.25 |
| | | 100 | 99.68 | 0.32 |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.4 µl at coverage factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmary statement not provided since not required.

Prepared by

....End of Calibration Certificate....

